# **Certificate of Analysis**



### TaqMan<sup>®</sup> Gene Expression Cells-to-CT<sup>™</sup> Kit

Product No. Lot No.	4399002 2826656
Date of Manufacture	2023-10-10
Expiration Date	2024-09-30
Product Description	The TaqMan <sup>®</sup> Gene Expression Cells-to-CT <sup>™</sup> Kit enables lysis and reverse transcription and real-time PCR (real-time RT-PCR) analysis of 10–10 <sup>5</sup> cultured cells, without isolating or purifying RNA.
Amount	40 Reactions
Functional Testing	Each component is functionally tested by preparing lysates from a serial dilution of 10–10 <sup>5</sup> cultured cells and performing real-time RT-PCR targeting PPIA. Amplification signals are compared to those obtained from purified RNA from the same source. <b>Result:</b> Pass

TEST	SPECIFICATION	RESULT
Lysis Solution		
Activity Assay Ionic Strength pH Assay	Activity of Lysis Solution is tested. Ionic strength is determined. pH is determined.	Pass Pass Pass
Stop Solution		
Exonuclease Activity	A sample is incubated with labeled double-stranded DNA, followed by PAGE analysis.	Pass
Inhibition Assay	The ability of Stop Solution to inhibit Lysis Solution activity is tested.	Pass
Nonspecific Endonuclease Activity	A sample is incubated with supercoiled plasmid DNA and analyzed by agarose gel electrophoresis.	Pass
pH Assay	pH is determined.	Pass
RNase Activity	A sample is incubated with labeled RNA, followed by PAGE analysis.	Pass
DNase I		
Protease Activity	A sample is incubated with protease substrate and analyzed by fluorescence.	Pass
RNase Activity	A sample is incubated with labeled RNA, followed by PAGE analysis.	Pass
2X RT Buffer		
Cation Concentration	Cation concentrations are determined.	Pass
Exonuclease Activity	A sample is incubated with labeled double-stranded DNA, followed by PAGE analysis.	Pass
Nonspecific Endonuclease Activity	A sample is incubated with supercoiled plasmid DNA and analyzed by agarose gel electrophoresis.	Pass

## **Certificate of Analysis**



pH Assay	pH is determined.	Pass		
RNase Activity	A sample is incubated with labeled RNA, followed by PAGE analysis.	Pass		
RT-PCR Testing	2X RT Buffer is tested in two-step RT-PCR targeting an endogenous control gene using purified RNA.	Pass		
20X RT Enzyme Mix				
Exonuclease Activity	A sample is incubated with labeled double-stranded DNA, followed by PAGE analysis.	Pass		
Nonspecific Endonuclease Activity	A sample is incubated with supercoiled plasmid DNA and analyzed by agarose gel electrophoresis.	Pass		
Protease Activity	A sample is incubated for with protease substrate and analyzed by fluorescence.	Pass		
RNAse Activity	A sample is incubated with labeled RNA, followed by PAGE analysis.	Pass		
RT Activity Assay	Incorporation of $[\alpha^{-32}P]$ dTTP into first strand cDNA during a reverse transcription reaction is measured.	Pass		
RT-PCR testing	20X RT Enzyme Mix is tested in two-step RT-PCR targeting an endogenous control gene using purified RNA.	Pass		
TaqMan <sup>®</sup> Gene Expression Master Mix				
AmpliTaq Gold <sup>®</sup> DNA Polymerase UP Activity	AmpliTaq Gold <sup>®</sup> DNA Polymerase Ultra-Pure (UP) activity is determined.	Pass		
Bacterial DNA	A sample is tested for bacterial DNA.	Pass		
dNTP Concentrations	Individual dNTP concentrations are measured.	Pass		
DNase Activity	A sample is tested for DNase activity.	Pass		
Mg <sup>2+</sup> Ion Concentration	A sample is tested for magnesium ion (Mg <sup>2+</sup> ) concentration.	Pass		
pH Assay	pH is determined.	Pass		
RNase Activity	A sample is tested for RNase activity.	Pass		
ROX <sup>™</sup> Passive Reference A sample is tested for ROX dye fluorescence. Pass Fluorescence Assay				
Functional Testing	TaqMan <sup>®</sup> Gene Expression Master Mix is tested with simplex and duplex PCR targeting plasmid DNAs on an Applied Biosystems 7900HT Fast Real-Time PCR System. Functional testing data for this component is available by entering its part number and lot number at http://docs.appliedbiosystems.com/coasearch.taf.	Pass		

For Research Use Only. Not for use in diagnostic procedures.

#### SO CERTIFICATION

This product is produced in accordance to Thermo Fisher Scientific Baltics UAB, a wholly owned subsidiary of Thermo Fisher Scientific Inc. (hereinafter Thermo Fisher) certified Quality Management (ISO 9001) and Environmental Management (ISO 14001) systems. Thermo Fisher Scientific Baltics UAB is also certified according to the standards ISO 13485 (Quality management system for IVD medical devices).

# **Certificate of Analysis**





J. Žilinskienė Quality Assurance Issued 2023-10-11