

CERTIFICATE OF ANALYSIS

60101 Dynabeads™ kilobaseBINDER™ Kit

Packaging Lot: 2901224

Expiry Date: 31.03.2026 (DD.MM.YYYY)

Storage: at 5±3°C

Filling lots for components in package:

| Lot | Quantity | Description |
|---------|----------|-------------------------------|
| 2859808 | 1 mL | Dynabeads™ M-280 Streptavidin |
| 2802725 | 15 mL | Washing Solution |
| 2855450 | 15 mL | Binding Solution |

QUALITY CONTROL

| Parameter | Method | Requirement | Result |
|------------------------|--|--|----------|
| Particle concentration | Dry weight measurement of beads. | 9.7-10.3 mg/mL | Conforms |
| Functional Assay | Tested for free biotin binding capacity. | 490 - 750 pmoles free biotin bound per mg coated beads | Conforms |

ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: **J. Žilinskienė**

