

## CERTIFICATE OF ANALYSIS

**60101 Dynabeads™ kilobaseBINDER™ Kit**

**Packaging Lot: 2977868**

**Expiry Date: 30.11.2026 (DD.MM.YYYY)**

**Storage: at 5±3°C**

### Filling lots for components in package:

Lot	Quantity	Description
2976858	1 mL	Dynabeads™ M-280 Streptavidin
2974779	15 mL	Washing Solution
2952488	15 mL	Binding Solution

### QUALITY CONTROL

Parameter	Method	Requirement	Result
Particle concentration	Dry weight measurement of beads.	9.7-10.3 mg/mL	Conforms
Functional Assay	Tested for free biotin binding capacity.	490 - 750 pmoles free biotin bound per mg coated beads	Conforms

### ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: **J. Žilinskienė**

