

CERTIFICATE OF ANALYSIS

60101 Dynabeads™ kilobaseBINDER™ Kit

Packaging Lot: 3052885

Expiry Date: 31.01.2027 (DD.MM.YYYY)

Storage: at 5±3°C

Filling lots for components in package:

Lot	Quantity	Description
3001149	1 mL	Dynabeads™ M-280 Streptavidin
3002147	15 mL	Washing Solution
2999891	15 mL	Binding Solution

QUALITY CONTROL

Parameter	Method	Requirement	Result
Particle concentration	Dry weight measurement of beads.	9.7-10.3 mg/mL	Conforms
Functional Assay	Tested for free biotin binding capacity.	490 - 750 pmoles free biotin bound per mg coated beads	Conforms

ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: **J. Žilinskienė**

