

CERTIFICATE OF ANALYSIS

783955000RX HT ExoSAP-IT High-Throughput PCR Product Cleanup

Packaging Lot: 2817924

Expiry Date: 31.10.2024 (DD.MM.YYYY)

Storage: at -20±5°C

Filling lots for components in package:

Lot	Quantity	Description
2718626	5 x 2 mL	HT ExoSAP-IT™ Reagent 1000 RX in tube

QUALITY CONTROL

Parameter	Method	Requirement	Result
Functional Assay	Test and Control ExoSAP-IT™ samples are used to eliminate unincorporated dNTP and unused primers from the PCR product. Exonuclease I/SAP-treated PCR product is sequenced. Quality of the sequence is evaluated visually.	There are no extra or missing bands, absence of stops in Test sequences. Test and Control sequences have uniform band intensity.	Conforms

ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: J. Žilinskienė

