

CERTIFICATE OF ANALYSIS

783955000RX HT ExoSAP-IT High-Throughput PCR Product Cleanup

Packaging Lot: 2955919

Expiry Date: 31.10.2025 (DD.MM.YYYY)

Storage: at -20±5°C

Filling lots for components in package:

Lot	Quantity	Description
2844750	5 × 2 mL	HT ExoSAP-IT™ Reagent 1000 RX in tube

QUALITY CONTROL

Parameter	Method	Requirement	Result
Functional Assay	ExoSAP-IT™ PCR Cleanup reagent is functionally tested to determine the activity of Exonuclease I and Shrimp Alkaline Phosphatase (SAP). Exonuclease I: Activity assay is performed in a primer degradation gel assay. SAP: Activity assay is performed in a p-nitrophenyl phosphate hydrolysis assay.	Meets predetermined specifications	Conforms

ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: J. Žilinskienė

