

CERTIFICATE OF ANALYSIS

785951EA HT ExoSAP-IT Fast High- Throughput PCR Product Cleanup

Packaging Lot: 2803077

Expiry Date: 30.09.2024 (DD.MM.YYYY)

Storage: at -20±5°C

Filling lots for components in package:

Lot	Quantity	Description
2801862	1 U	HT ExoSAP-IT™ Fast Reagent

QUALITY CONTROL

Parameter	Method	Requirement	Result
Functional Assay	Test and Control HT ExoSAP-IT™ Fast samples are used to eliminate unincorporated dNTP and unused primers from the PCR product. Exonuclease I/SAP-treated PCR product is sequenced. Quality of the sequence is evaluated visually.	There are no extra or missing bands, absence of Stops in Test sequences. Test and Control sequences have uniform band intensity.	Conforms

ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: **J. Žilinskienė**

