

Certificate of Analysis

TaqPath[™] 1-Step Multiplex Master Mix (No ROX)

Product No: \square A28521 \square A28522 \square A28523Quantity: \square 1 x 0.5 mL \square 5 x 1 mL \square 1 x 10 mLPackaging Lot¹:2794510Expiration Date:04 July 2025Storage:-30 to -10°CBulk Lot²:2757325

Manufactured according to current good manufacturing practices (cGMP) and subjected to a panel of quality control tests to ensure the highest level of performance and lot-to-lot consistency.
Analytical Methods

Parameter QC method Specification Result Magnesium ion (Mg²⁺) High performance ion chromatography with 24.48 - 33.12 mM Conforms concentration the detection of conductivity signal is Potassium ion (K+) performed on the product and standards. 170 – 230 mM Conforms concentration DNase levels are measured using a modified DNA **DNase Level** oligonucleotide possessing a Samples have $\leq 44.0 \text{ pg} / 5 \mu \text{L}$ Conforms quencher and fluorescent label that emits a fluorescent signal when cleaved by DNase. E. coli DNA levels are quantified via gPCR ≤ 10 copies of E. coli DNA / 50 µL E.coli DNA Level utilizing a standard curve with four replicates Conforms **Reaction Volume** for each point. The pH value is measured at a temperature 7.95 - 8.25 pH Conforms of 25 °C.

Note: Enzyme bulk is tested for activity and E. coli DNA Level

Functional Tests

TaqPath[™] 1-Step Multiplex Master Mix (No ROX) is tested for performance using a panel of two quadplex assay combinations and four simplex Gene Expressions assays. The first quadplex assay combination consists of four Gene Expression assays run with a serial dilution of human RNA.. The second quadplex assay combination consists of three Gene Expression assays coupled with an exogenous internal positive control (IPC) reagent. The second quadplex and four simplex reactions are run at one concentration of purified human RNA. Performance attributes evaluated include: PCR efficiency, R2 value, Ct Value, and Absolute Delta Ct between Quadplex and Simplex.

Quadplex 1	PCR Efficiency	R ² Value	Result
Assay Hs00164924_m1	85% - 115%	≥ 0.95	Conforms
Assay Hs99999909_m1	85% - 115%	≥ 0.95	Conforms
Assay RNaseP	85% - 115%	≥ 0.95	Conforms
Assays GAPDH	85% - 115%	≥ 0.95	Conforms

www.thermofisher.com

Thermo Fisher Scientific • 7335 Executive Way • Frederick • MD 21704 • Tel. 800.955.6288 • Email: techservices@thermofisher.com

Thermo Fisher SCIENTIFIC

Certificate of Analysis

Quadplex 2	Average Ct Value	Ct Std. Dev.	Result
Assay Exogenous IPC (quadplex assay)	27.18 - 31.48	≤ 0.4	Conforms

Simplex	Absolute Delta Ct (Simplex vs Multiplex)	Result
Assay Hs00164924_m1	< 1.5	Conforms
Assay Hs99999909_m1	< 1.5	Conforms
Assay RNaseP	< 1.5	Conforms
Assays GAPDH	< 1.5	Conforms



Quality Signature: Audronė Lakštauskienė

Date: 1 September 2023

For Laboratory Use.

¹ Packaging Lot is the unique Lot # assigned to the packing event of the vials into the product boxes.

² Bulk Lot is the unique Lot # assigned to the production Lot prior to filling. Note: It is possible for different packing Lot's to be filled by the same bulk Lot.

www.thermofisher.com

Thermo Fisher Scientific • 7335 Executive Way • Frederick • MD 21704 • Tel. 800.955.6288 • Email: techservices@thermofisher.com