

CERTIFICATE OF ANALYSIS

A36582 MagMAX™ DNA Multi-Sample Ultra 2.0 Elution Solution

Packaging Lot: 2850674

Expiry Date: 30.11.2025 (DD.MM.YYYY)

Storage: at room temperature

Filling lots for components in package:

Lot	Quantity	Description
2850674	120 mL	MagMAX DNA Multi-Sample Ultra 2.0 Elution Solution

QUALITY CONTROL

Parameter	Method	Requirement	Result
Exonuclease activity	A sample is incubated with labeled double-stranded DNA, followed by PAGE analysis.	≤ 30% as of control	Conforms
Nonspecific Endonuclease activity	A sample is incubated with supercoiled plasmid DNA and analyzed by agarose gel electrophoresis.	≤ 30% as of control	Conforms
pH	Determined according to Ph. Eur. 2.2.3.	Solution specific	Conforms
Conductivity	Determined according to Ph. Eur. at 25 °C.	Solution specific	Conforms
Human DNA Contamination	Contamination determined by RT-PCR.	Solution specific	Conforms

ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: **J. Žilinskienė**

