

## CERTIFICATE OF ANALYSIS

### A43190024 Collibri™ DNA Library Cleanup Kit

Packaging Lot: 2808541

Expiry Date: 30.04.2024 (DD.MM.YYYY)

Storage: at 5±3°C

#### Filling lots for components in package:

Lot	Quantity	Description
2807350	10 mL	DNA Cleanup Beads
2707194	5 mL	Elution Buffer
2719569	4.5 mL	Wash Buffer (Concentrated)

#### QUALITY CONTROL

Parameter	Method	Requirement	Result
Functional test	Cleanup of DNA fragment mixture and downstream assessment of library size and yield with Agilent 2100 Bioanalyzer instrument using High Sensitivity DNA Kit (Agilent Technologies Inc.)	Removal of 50 and 150 bp fragments and ≥50% yield recovery of 500 and ≥80% yield of 600-1000 bp DNA fragments.	Conforms

#### ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: **J. Žiliuskienė** 