

## CERTIFICATE OF ANALYSIS

**AM10102**      **BrightStar™-Plus Positively Charged Nylon Membrane**

**Packaging Lot:**      **2856658**

**Manufacturing Date:**      21.11.2023 (DD.MM.YYYY)

**Storage:**      at room temperature

### QUALITY CONTROL

Parameter	Method	Requirement	Result
Material	Material specification.	Positively charged, supported nylon 6.6	Conforms
Pore Size	Measurement of pore size.	0.45 µm	Conforms
Thickness	Measurement of material thickness.	5.5 – 6.7 mils	Conforms

#### OTHER INFORMATION

##### **Limited Product Warranty**

Thermo Fisher Scientific Inc. and/or its affiliate(s) warrant their products as set forth in the Life Technologies General Terms and Conditions of Sale found on Thermo Fisher Scientific website at [www.thermofisher.com/termsandconditions](http://www.thermofisher.com/termsandconditions).

#### ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: **J. Žilinskienė**

