

## CERTIFICATE OF ANALYSIS

**AM8670**      **ULTRAhyb™ Ultrasensitive Hybridization Buffer**

**Packaging Lot:** 2969329

**Expiry Date:** 30.04.2028 (DD.MM.YYYY)

**Storage:** at 5±3°C

### QUALITY CONTROL

| Parameter           | Method  | Requirement | Result   |
|---------------------|---|-------------|----------|
| pH                  | Determined according to Ph. Eur. 2.2.3.   | 7.5 ± 0.1   | Conforms |
| RNase Contamination | No detectable RNase contamination was observed using fluorimetric (RNase A equivalents) test. | Conforms    | Conforms |
| DNase Contamination | No detectable DNase contamination was observed using fluorimetric (DNase I equivalents) test. | Conforms    | Conforms |

#### ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: **J. Žilinskienė**

