

Certificate of Analysis

ThermoFisher
SCIENTIFIC

0.5 M EDTA, pH 8.0

Product No. **AM9260G**
Lot No. **2825019**
Date of Manufacture **2023-10-10**

TEST	SPECIFICATION	RESULT
Nonspecific endonuclease activity		
A sample is incubated with supercoiled plasmid DNA and analyzed by agarose gel electrophoresis.	≤30% change in probe relative to control	Pass
Exonuclease activity		
A sample is incubated with labeled double stranded DNA, followed by PAGE analysis.	≤30% change in probe relative to control	Pass
RNase activity		
A sample is incubated with labeled RNA and analyzed by PAGE.	≤30% change in probe relative to control	Pass

OTHER INFORMATION

Limited Product Warranty Thermo Fisher Scientific Inc. and/or its affiliate(s) warrant their products as set forth in the Life Technologies' General Terms and Conditions of Sale found on Thermo Fisher Scientific website at www.thermofisher.com/termsandconditions.

For Research Use Only. Not for use in diagnostic procedures.

ISO CERTIFICATION

This product is produced in accordance to Thermo Fisher Scientific Baltics UAB, a wholly owned subsidiary of Thermo Fisher Scientific Inc. (hereinafter Thermo Fisher) certified Quality Management (ISO 9001) and Environmental Management (ISO 14001) systems. Thermo Fisher Scientific Baltics UAB is also certified according to the standards ISO 13485 (Quality management system for IVD medical devices).


Thermo Fisher Scientific Baltic UAB

V.A. Graiciuno 8

Vilnius, Lithuania 02241

www.thermofisher.com

For inquiries, contact us at cofarequests@thermofisher.com


J. Žilinskienė
Quality Assurance
Issued 2023-10-11