

CERTIFICATE OF ANALYSIS

K210016 PureLink HiPure Plasmid Filter Maxiprep Kit

Packaging Lot: 2823095

Expiry Date: 30.09.2025 (DD.MM.YYYY)

Storage: at room temperature

Filling lots for components in package:

Lot	Quantity	Description
2812585	0.65 mL	RNase, 20 mg/ml
2814628	30 mL	TE Buffer (TE)
2814765	110 mL	Resuspension Buffer (R3)
2813810	110 mL	Lysis Buffer (L7)
2815668	110 mL	Precipitation Buffer (N3)
2812666	175 mL	Elution Buffer (E4)
2812811	330 mL	Equilibration Buffer (EQ1)
2822498	2 x 330 mL	Wash Buffer (W8)
2811091	1 pack	Column Holder maxi
2818477	2 x 1 pack	HiPure Maxi Filter Columns

QUALITY CONTROL

Parameter	Method	Requirement	Result
Yield of purified control plasmid DNA	PureLink HiPure Plasmid Filter Maxiprep Kit is qualified by isolating a high copy number control plasmid DNA from 100 ml of overnight E. coli culture following the protocol outlined in the manual.	$\geq 750 \mu\text{g}$	Conforms
Purity	A260/A280 ratio.	≥ 1.8	Conforms
Flow rate	Average accumulated column processing time.	30 - 45 min	Conforms
Gel electrophoresis	1 μg of purified plasmid DNA is run on an agarose gel.	A dominant band of supercoiled plasmid DNA is observed, no RNA visible	Conforms
Digestion	The plasmid DNA is evaluated by digestion (1 μg 15 min) of purified plasmid DNA with Anza™11 EcoRI restriction endonuclease.	DNA - completely digested, no additional fragments	Conforms

ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by QC: **J. Žilinskienė**

