

CERTIFICATE OF ANALYSIS

N8080240 AmpliTaq Gold™ DNA Polymerase with Buffer I, 250 units

Packaging Lot: 2795927

Expire Date: 06.2025

Filling lots for components in package:

Lot	Description
2751316	AmpliTaq Gold DNA Polymerase, 250 units
2738663	10X PCR Buffer I

QUALITY CONTROL

TEST	SPECIFICATION	RESULT
AMPLITAQ GOLD DNA POLYMERASE		
ANALYTICAL TEST		
Polymerase activity	5.0 - 6.5 U/µL	Pass
Purity (HPLC - UV detection)	≥ 95.0 %	Pass
Purity (HPLC - fluorimetric detection)	≥ 95.0 %	Pass
Purity (SDS PAGE)	≥ 95.0 %	Pass
Amplification efficiency of human genomic DNA	85 - 105 %	Pass
Human DNA	≤ 2 copies / 10 U	Pass
Bacterial DNA	≤ 1 copy E.coli DNA / 5 U	Pass
ssDNA DNA nuclease (gel-based)	None detected / 40 U	Pass
dsDNA DNA nuclease (gel-based)	None detected / 40 U	Pass
RNase A equivalents (fluorimetric assay)	≤ 1.15 pg / 1 U	Pass
DNase I equivalents (fluorimetric assay)	≤ 155 pg / 1 U	Pass
10X PCR BUFFER I		
ANALYTICAL TEST		
pH at 25°C	8.15 – 8.30	Pass
K ⁺ concentration (HPIC)	425.0 – 575.0 mM	Pass
Mg ²⁺ concentration (HPIC)	13.50 – 17.00 mM	Pass

Amplification of human genomic DNA

Specific band on gel

Pass

For Research Use Only. Not for use in diagnostic procedures.

ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management systems.

Quality authorized by QC: **J. Žiliinskienė**

