



Ansell Healthcare, LLC
111 Wood Avenue South, Suite 210, Iselin, NJ 08830

CERTIFICATE *of* CONFORMANCE

Product: Nitrilite Silky
Style: 93-311
10-Digit Number: 18123118EV

Property	Ansell Specification	Test Method	Results
Particles Count 0.5-20u	< 2400 particles/cm ²	IENT-RP-CC005.3	CONFORMS
Chloride	< 0.600 ug/cm ²	IENT-RP-CC005.3	CONFORMS
Nitrate	< 0.500 ug/cm ²	IENT-RP-CC005.3	CONFORMS
Sulphate	< 0.200 ug/cm ²	IENT-RP-CC005.3	CONFORMS
Sodium	< 0.500 ug/cm ²	IENT-RP-CC005.3	CONFORMS
Potassium	< 0.050 ug/cm ²	IENT-RP-CC005.3	CONFORMS
Lithium	< 0.0004 ug/cm ²	IENT-RP-CC005.3	CONFORMS
Magnesium	< 0.050 ug/cm ²	IENT-RP-CC005.3	CONFORMS
Calcium	< 0.500 ug/cm ²	IENT-RP-CC005.3	CONFORMS
Zinc	< 0.300 ug/cm ²	IENT-RP-CC005.3	CONFORMS
Aluminum	< 0.050 ug/cm ²	IENT-RP-CC005.3	CONFORMS
Iron	< 0.050 ug/cm ²	IENT-RP-CC005.3	CONFORMS
Copper	< 0.050 ug/cm ²	IENT-RP-CC005.3	CONFORMS
Non Volatile Residue	< 5.0 ug/cm ²	IENT-RP-CC005.3	CONFORMS
Silicone Oil, Amide and DOP	Absent	IENT-RP-CC005.3	CONFORMS
Electrostatic Decay (50% RH)	< 5.0 seconds	ASTM D257, EST S11.11	CONFORMS
Surface Resistivity (50% RH)	1 x 10 ¹⁰ ohms/square	ASTM D257, EST S11.11	CONFORMS
Grip Coefficient-Donning	< 90 gram-force	Ansell Test Method DE//009	CONFORMS
Grip Coefficient-Gripping	110 < x < 220 gram-force	Ansell Test Method DE//009	CONFORMS
Pin Holes	AQL 2.5; G1	ASTM D 5151	CONFORMS

Certified by

A handwritten signature in black ink, appearing to read "Ambir Hukkanen". The signature is fluid and cursive, with a long horizontal stroke extending to the right.

Ambir Hukkanen ASQ CMQ/OE, CQA
Associate Director Quality Assurance

Date:
