

CERTIFICATE OF CONFORMITY & CERTIFICATE OF ANALYSIS

Brand	AccuTech®
Size(s)	6
Material	Natural Rubber Latex
Style	91-225
Short Description	Sterile Disposable Latex Cleanroom Glove
Lot Number	250204
Batch Number	2502046804
Manufacturing Date	2025-02
Expiration Date	2030-01

CERTIFICATE OF CONFORMITY

Ansell certifies that the above product is manufactured in accordance with Ansell product specifications and is in conformity with the following standards or normative documents:

Standard/Normative Document	Result
FDA 21 CFR 177-2600	Conforms
Free from BSE/TSE (EMA/410/01)	Conforms
ISO 11137-1	Conforms
PPER (EU) 2016/425	Conforms
Reach Regulation	Conforms

CERTIFICATE OF ANALYSIS

Property	Test Method	Specification	Result
Freedom from Holes - Final Release	EN455-1/ASTM D3577	G1, AQL 0.65	Conforms
Dimension Test	ASTM D3767	S-2, AQL 4.0	Conforms
Glove Length (mm)	ASTM D3767	Min. 285	Conforms
Palm Thickness (mm)	ASTM D3767	0.19 – 0.23	Conforms
Physical Properties	ASTM D412	S-2, AQL 4.0	Conforms
Tensile Strength (MPa)	ASTM D412	Min. 24	Conforms
Ultimate Elongation (%)	ASTM D412	Min. 750	Conforms
Force at Break (N)	EN 455-2	Min. 9	Conforms
Protein Content (µg/g)	EN 455-3	<50.0	Conforms

Electronically Signed by:

Amy Jones

Manager, QA

Date: 29 April 2025



For additional information visit us at www.ansell.com, or call us at:

Europe, Middle East & Africa Region

Ansell Healthcare Europe N.V.

T: +32 (0) 2 528 74 00

F: +32 (0) 2 528 74 01

Australia

Ansell Limited

T: +61 1800 337 041

F: +61 1800 803 578

Latin America & Caribbean Region

Ansell Commercial Mexico S.A. de C.V.

T: +52 442 248 1544 / 248 3133

North America Region

Ansell Healthcare products LLC

T: +1 800 800 0444

F: +1 800 800 0445

Asia Pacific Region

Ansell Global Trading Center

(Malaysia) Sdn Bhd

T: +603 8310 6688

F: +603 8310 6699