



## CERTIFICATE OF CONFORMITY & CERTIFICATE OF ANALYSIS

|                    |                                          |
|--------------------|------------------------------------------|
| Brand              | AccuTech®                                |
| Size(s)            | 9                                        |
| Material           | Natural Rubber Latex                     |
| Style              | 91-225                                   |
| Short Description  | Sterile Disposable Latex Cleanroom Glove |
| Lot Number         | 240504                                   |
| Batch Number       | 2405072204                               |
| Manufacturing Date | 2024-05                                  |
| Expiration Date    | 2029-04                                  |

### DECLARATION OF CONFORMITY

Ansell certifies that the above product is manufactured in accordance with Ansell product specifications and is in conformity with the following standards or normative documents:

| Standard/Normative Document    | Result   |
|--------------------------------|----------|
| FDA 21 CFR 177-2600            | Conforms |
| Free from BSE/TSE (EMA/410/01) | Conforms |
| ISO 11137-1                    | Conforms |
| PPER (EU) 2016/425             | Conforms |
| Reach Regulation               | Conforms |

### DECLARATION OF ANALYSIS

| Property                           | Test Method        | Specification | Result   |
|------------------------------------|--------------------|---------------|----------|
| Freedom from Holes - Final Release | EN455-1/ASTM D3577 | G1, AQL 0.65  | Conforms |
| Dimension Test                     | ASTM D3767         | S-2, AQL 4.0  | Conforms |
| Glove Length (mm)                  | ASTM D3767         | Min. 285      | Conforms |
| Palm Thickness (mm)                | ASTM D3767         | 0.19 – 0.23   | Conforms |
| Physical Properties                | ASTM D412          | S-2, AQL 4.0  | Conforms |
| Tensile Strength (MPa)             | ASTM D412          | Min. 24       | Conforms |
| Ultimate Elongation (%)            | ASTM D412          | Min. 750      | Conforms |
| Force at Break (N)                 | EN 455-2           | Min. 9        | Conforms |
| Protein Content (µg/g)             | EN 455-3           | <50.0         | Conforms |

Electronically Signed by:  
**Amy Jones**  
Manager, QA

Date: 21 August 2024



For additional information visit us at [www.ansell.com](http://www.ansell.com), or call us at:

**Europe, Middle East & Africa Region**

Ansell Healthcare Europe N.V.

T: +32 (0) 2 528 74 00

F: +32 (0) 2 528 74 01

**Australia**

Ansell Limited

T: +61 1800 337 041

F: +61 1800 803 578

**Latin America & Caribbean Region**

Ansell Commercial Mexico S.A. de C.V.

T: +52 442 248 1544 / 248 3133

**North America Region**

Ansell Healthcare products LLC

T: +1 800 800 0444

F: +1 800 800 0445

**Asia Pacific Region**

Ansell Global Trading Center

(Malaysia) Sdn Bhd

T: +603 8310 6688

F: +603 8310 6699