



CERTIFICATE OF IRRADIATION

Brand	TouchNTuff®
Size(s)	XL (9.5 - 10)
Material	Nitrile
Style	93-700
Short Description	Sterile Disposable Nitrile Cleanroom Glove
Lot Number	2202EV
Batch Number	22027002EV
Manufacturing Date	2022-02
Expiration Date	2025-01

Ansell certifies that the product above has been successfully sterilized as follows in accordance with Ansell product specifications:

Min. Specified Irradiation Dose (kGy)	25
Irradiation Location	Synergy Sterilisation Rawang (M) Sdn Bhd, Lot 42, Rawang Integrated Industrial Park, 48000 Selangor Darul Ehsan, Malaysia
Irradiation Lot No	S1265057611
Irradiation Date	27 February 2022
Min. Dose Delivered (kGy)	28.0
Max. Dose Delivered (kGy)	36.5

Electronically Signed by:

Amy Jones

Manager, QA

Date: 8 August 2024

For additional information visit us at www.ansell.com, or call us at:

Europe, Middle East & Africa Region

Ansell Healthcare Europe N.V.

T: +32 (0) 2 528 74 00

F: +32 (0) 2 528 74 01

Australia

Ansell Limited

T: +61 1800 337 041

F: +61 1800 803 578

Latin America & Caribbean Region

Ansell Commercial Mexico S.A. de C.V.

T: +52 442 248 1544 / 248 3133

North America Region

Ansell Healthcare products LLC

T: +1 800 800 0444

F: +1 800 800 0445

Asia Pacific Region

Ansell Global Trading Center

(Malaysia) Sdn Bhd

T: +603 8310 6688

F: +603 8310 6699