

CERTIFICATE OF IRRADIATION

Brand	AccuTech®
Size(s)	8
Material	Natural Rubber Latex
Style	91-225
Short Description	Sterile Disposable Latex Cleanroom Glove
Lot Number	241104
Batch Number	2411104804
Manufacturing Date	2024-11
Expiration Date	2029-10

Ansell certifies that the product above has been successfully sterilized as follows in accordance with Ansell product specifications:

Min. Specified Irradiation Dose (kGy)	25
Irradiation Location	Ansell N.P. Sdn Bhd, Lot 80, Air Keroh Industrial Estate, Melaka, 75450 Air Keroh, Malaysia
Irradiation Lot No	2411104804
Irradiation Date	15 January 2025
Min. Dose Delivered (kGy)	26.20
Max. Dose Delivered (kGy)	35.20

Electronically Signed by:

Amy Jones

Manager, QA

Date: 14 February 2025

For additional information visit us at www.ansell.com, or call us at:

Europe, Middle East & Africa Region

Ansell Healthcare Europe N.V.

T: +32 (0) 2 528 74 00

F: +32 (0) 2 528 74 01

Australia

Ansell Limited

T: +61 1800 337 041

F: +61 1800 803 578

Latin America & Caribbean Region

Ansell Commercial Mexico S.A. de C.V.

T: +52 442 248 1544 / 248 3133

North America Region

Ansell Healthcare products LLC

T: +1 800 800 0444

F: +1 800 800 0445

Asia Pacific Region

Ansell Global Trading Center

(Malaysia) Sdn Bhd

T: +603 8310 6688

F: +603 8310 6699