

CERTIFICATE OF IRRADIATION

Brand	AccuTech®
Size(s)	8
Material	Natural Rubber Latex
Style	91-225
Short Description	Sterile Disposable Latex Cleanroom Glove
Lot Number	250204
Batch Number	2502048904
Manufacturing Date	2025-02
Expiration Date	2030-01

Ansell certifies that the product above has been successfully sterilized as follows in accordance with Ansell product specifications:

Min. Specified Irradiation Dose (kGy)	25
Irradiation Location	Ansell N.P. Sdn Bhd, Lot 80, Air Keroh Industrial Estate, Melaka, 75450 Air Keroh, Malaysia
Irradiation Lot No	2502048904
Irradiation Date	10 April 2025
Min. Dose Delivered (kGy)	26.40
Max. Dose Delivered (kGy)	34.90

Electronically Signed by:

Amy Jones

Manager, QA

Date: 3 June 2025

For additional information visit us at www.ansell.com, or contact us at:

Europe, Middle East & Africa Region

Ansell Healthcare Europe N.V.

Riverside Business Park

Bldv International, 55

B-1070 Brussels,

Belgium

T: +32 (0) 2 528 74 00

F: +32 (0) 2 528 74 01

Australia & Asia Pacific Region

678 Victoria St,

Richmond VIC 3121

Australia

Tel: +61 1800 337 041

Fax: +611 800 803578

Email: anzcustomerservice@ansell.com

North America Region

Ansell Healthcare products LLC

111 S Wood Ave,

Suite 900 Iselin,

NJ 08830,

United States

T: +1 800 800 0444

F: +1 800 800 0445

CA Tel: +1 800 363 8340

Email: customerserviceus@ansell.com

Ansell Healthcare Products LLC

2301 Robb Drive,

Reno,

NV 89523,

United States

US Tel: +1 800 800 0444

US Fax: +1 800 800 0445

Email: customerserviceus@ansell.com

Latin America & Caribbean Region

Ansell Comercial México S.A. de C.V.

Blvd. Bernardo Quintana 7001-C,

Q7001 Torre 2, Suites 1304-1306 Centro Sur.

Querétaro, Qro. México C.P. 76090

Tel: +52 442 296 2050

Email: cslac@ansell.com

Ansell Brazil Ltda.

Rua das Figueiras, 474

4º andar - Bairro Jardim

Santo André - SP

CEP: 09080-300 – Brasil

Tel: +55 11 3356 3100

Email: sac@ansell.com