

CERTIFICATE OF IRRADIATION

Brand	AccuTech®
Size(s)	6.5
Material	Natural Rubber Latex
Style	91-225
Short Description	Sterile Disposable Latex Cleanroom Glove
Lot Number	250504
Batch Number	2505032604
Manufacturing Date	2025-05
Expiration Date	2030-04

Ansell certifies that the product above has been successfully sterilized as follows in accordance with Ansell product specifications:

Min. Specified Irradiation Dose (kGy)	25
Irradiation Location	Ansell N.P. Sdn Bhd, Lot 80, Air Keroh Industrial Estate, Melaka, 75450 Air Keroh, Malaysia
Irradiation Lot No	2505032604
Irradiation Date	18 June 2025
Min. Dose Delivered (kGy)	26.10
Max. Dose Delivered (kGy)	34.90

Electronically Signed by:

Amy Jones

Sr. Manager, QA

Reno, NV, United States

Date: 14 August 2025



For additional information visit us at www.ansell.com, or contact us at:

Europe, Middle East & Africa Region

Ansell Healthcare Europe N.V.
Riverside Business Park
Blvd International, 55
B-1070 Brussels,
Belgium
T: +32 (0) 2 528 74 00
F: +32 (0) 2 528 74 01

Australia & Asia Pacific Region

678 Victoria St,
Richmond VIC 3121
Australia
Tel: +61 1800 337 041
Fax: +611 800 803578
Email: anzcustomerservice@ansell.com

North America Region

Ansell Healthcare products LLC
111 S Wood Ave,
Suite 900 Iselin,
NJ 08830,
United States
T: +1 800 800 0444
F: +1 800 800 0445
CA Tel: +1 800 363 8340
Email: customerserviceus@ansell.com

Ansell Healthcare Products LLC
2301 Robb Drive,
Reno,
NV 89523,
United States
US Tel: +1 800 800 0444
US Fax: +1 800 800 0445
Email: customerserviceus@ansell.com

Latin America & Caribbean Region

Ansell Comercial México S.A. de C.V.
Blvd. Bernardo Quintana 7001-C,
Q7001 Torre 2, Suites 1304-1306 Centro Sur.
Querétaro, Qro. México C.P. 76090
Tel: +52 442 296 2050
Email: cslac@ansell.com

Ansell Brazil Ltda.

Rua das Figueiras, 474
4º andar - Bairro Jardim
Santo André - SP
CEP: 09080-300 – Brasil
Tel: +55 11 3356 3100
Email: sac@ansell.com