

CERTIFICATE OF IRRADIATION

Brand	TouchNTuff®
Size(s)	M (7.5 - 8), L (8.5 - 9)
Material	Nitrile
Style	93-700
Short Description	Sterile Disposable Nitrile Cleanroom Glove
Lot Number	2504SW
Batch Number	25047002SW
Manufacturing Date	2025-04
Expiration Date	2028-03

Ansell certifies that the product above has been successfully sterilized as follows in accordance with Ansell product specifications:

Min. Specified Irradiation Dose (kGy)	25
Irradiation Location	Synergy Sterilisation Rawang (M) Sdn Bhd, Lot 42, Rawang Integrated Industrial Park, 48000 Selangor Darul Ehsan, Malaysia
Irradiation Lot No	3043-3037A
Irradiation Date	19 April 2025
Min. Dose Delivered (kGy)	28.2
Max. Dose Delivered (kGy)	37.2

Electronically Signed by:

Amy Jones

Manager, QA

Date: 14 May 2025

For additional information visit us at www.ansell.com, or call us at:

Europe, Middle East & Africa Region

Ansell Healthcare Europe N.V.

T: +32 (0) 2 528 74 00

F: +32 (0) 2 528 74 01

Australia

Ansell Limited

T: +61 1800 337 041

F: +61 1800 803 578

Latin America & Caribbean Region

Ansell Commercial Mexico S.A. de C.V.

T: +52 442 248 1544 / 248 3133

North America Region

Ansell Healthcare products LLC

T: +1 800 800 0444

F: +1 800 800 0445

Asia Pacific Region

Ansell Global Trading Center

(Malaysia) Sdn Bhd

T: +603 8310 6688

F: +603 8310 6699