

Test Report

– Eppendorf Sterile –

1 Set epT.I.P.S.

2-200 µL

Sterile, individually wrapped



2027-03

Eppendorf AG
22331 Hamburg · Germany
Tel: +49 1803 666 789
E-Mail: eppendorf@eppendorf.com

LOT L2030811

STERILE R

Order no.: 0030 027.035

Distributed in North America by
Eppendorf North America · USA
Tel: +1 800 645 3050
E-Mail: info@eppendorf.com

Made in Germany

Cat. no.: 022492209

eppendorf

• **Sterility**

Method:

Sterilization with SAL 10⁻⁶ by  Irradiation according DIN EN ISO 11137:2015* or
 ethylene oxide*

*please see attached label for the standard method of sterilization for this product

Test method:

USP, Ph. Eur. 8.0 / 2.6.12

• **Endotoxin**

Test method:

Kinetic-chromogenic LAL test (Ph. Eur. 8.0 / 2.6.14)

Limit:

< 0.001 EU/mL

Customer: Eppendorf AG, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: **26.05.22**

Testing of the above described lot showed conformity within the limits of detection.

23.05.22

Date of release

Epp 2205 1083

Registration number

Th. Weigel

Dipl.-Biol. Thomas Weigel (Head)

The test results refer exclusively to the items tested. The certificate must not be copied partially without the approval of LADR GmbH.