



Certificate of Conformance, Analysis, and Sterility

This certification is provided as full assurance that the following product code and lot number were manufactured in accordance via a Quality System certified to ISO 9001 with prescribed procedures and specifications.

Product Code:	PS-911EB/ETOH
Lot Number:	1113283-0-0-1
Manufacturing Date:	01/31/2025
Ethyl Alcohol(denatured):	70 ± 3% by volume with the remainder DI H2O.
Expiration Date:	01/2027

Analysis

Quantity of Particles Po:	1.772 millions/m ²
---------------------------	-------------------------------

Sterility

Gamma Processing Run ID:	3701903
Irradiation Dose Specified:	18.2 kGy to 40.0 kGy
Irradiation Dose Delivered:	19.0 kGy to 34.9 kGy
Validated Sterile Product:	Validation Sterile packets available upon request.
Sterility Assurance Level:	10 ⁻⁶ (SAL determined by test method AAMI/ISO 11137)

QC Supervisor
coc@contecinc.com

Thursday, February 13, 2025

NORTH AMERICA

Contec, Inc.
525 Locust Grove
Spartanburg, SC 29303 USA
tel: +1 864 503 8333
fax: +1 864 503 8444

EUROPE

Contec Europe
ZI Du Prat, Avenue Paul Duplex
56000 VANNES France
tel: +33 (0)2 97 43 76 98
fax: +33 (0) 97 43 76 86

ASIA

Contec Cleanroom Technology (Suzhou) Co., Ltd.
17 Longyun Road
Suzhou Industrial Park, Suzhou 215024 China
tel: +86 512 6274 4050
fax: +86 512 6274 4051

web: www.contecinc.com
e-mail: info@contecinc.com



Certificate of Processing

STERIGENICS 10811 Withers Cove Park Dr. Charlotte NC 28278
TEL 704 588-6877 FAX 704 588-8355 www.sterigenics.com

R55480102

02/12/2025 17:43:05 GMT
Page 1 of 1

Customer Name: Contec- PO Box 530
P.O.# 147274

Processing Facility:

Charlotte Gamma

Work Order # 3701903
Sales Order # 3734344

18.2-40.0 kGy

CNT4, Gamma Treatment

Irradiation Date/Time:

02/11/2025 13:32:00 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	512	CA	PS-911EB/ETOH	Clean Room Product	1113283-0-0-1	147274
	512	CA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By-----	
						User	Date /Time
450.00	Minimum Dose	18.2 kGy	40.0 kGy	19.0 KGY	Pass	TDIAZ	02/12/2025 03:25:14 GMT
		Reason Code Test				Faustino (Tino) Diaz	
450.00	Maximum Dose	18.2 kGy	40.0 kGy	34.9 KGY	Pass	TDIAZ	02/12/2025 03:25:36 GMT
		Reason Code Test				Faustino (Tino) Diaz	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: WENDY KOBLISH
Reason: Work Order Completions

Date: 02/12/2025 17:33:36 GMT