



## Certificate of Conformance, Analysis, and Sterility

This certification is provided as full assurance that the following product code and lot number were manufactured in accordance via a Quality System certified to ISO 9001 with prescribed procedures and specifications.

---

Product Code:	PS-911EB/ETOH
Lot Number:	1133273-0-0-1
Manufacturing Date:	04/25/2025
Ethyl Alcohol(denatured):	70 ± 3% by volume with the remainder DI H2O.
Expiration Date:	04/2027

---

### Analysis

Quantity of Particles Po:	1.535 millions/m <sup>2</sup>
---------------------------	-------------------------------

---

### Sterility

Gamma Processing Run ID:	3753176
Irradiation Dose Specified:	18.2 kGy to 40.0 kGy
Irradiation Dose Delivered:	22.4 kGy to 34.4 kGy
Validated Sterile Product:	Validation Sterile packets available upon request.
Sterility Assurance Level:	10 <sup>-6</sup> (SAL determined by test method AAMI/ISO 11137)

---

QC Supervisor  
coc@contecinc.com

Monday, May 05, 2025

#### NORTH AMERICA

Contec, Inc.  
525 Locust Grove  
Spartanburg, SC 29303 USA  
tel: +1 864 503 8333  
fax: +1 864 503 8444

#### EUROPE

Contec Europe  
ZI Du Prat, Avenue Paul Dupleix  
56000 VANNES France  
tel: +33 (0)2 97 43 76 98  
fax: +33 (0) 97 43 76 86

#### ASIA

Contec Cleanroom Technology (Suzhou) Co., Ltd.  
17 Longyun Road  
Suzhou Industrial Park, Suzhou 215024 China  
tel: +86 512 6274 4050  
fax: +86 512 6274 4051

web: [www.contecinc.com](http://www.contecinc.com)  
e-mail: [info@contecinc.com](mailto:info@contecinc.com)

Page 1 of 1

DID: 2061B

## Certificate of Processing

STERIGENICS 10811 Withers Cove Park Dr. Charlotte NC 28278  
 TEL 704 588-6877 FAX 704 588-8355 [www.sterigenics.com](http://www.sterigenics.com)

R55480102

05/02/2025 13:05:33 GMT  
 Page 1 of 1

<b>Customer Name:</b>	Contec- PO Box 530	<b>Processing Facility:</b>	Charlotte Gamma	<b>Work Order #</b>	3753176
<b>P.O.#</b>	149470			<b>Sales Order #</b>	3795995
	18.2-40.0 kGy	CNT4, Gamma Treatment	<b>Irradiation Date/Time:</b>	05/01/2025 08:19:00 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	414	CA	PS-911EB/ETOH	Clean Room Product	1133273-0-0-1	149470
	414	CA	<b>Total</b>			

### Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	18.2 kGy	40.0 kGy	22.4 KGY	Pass	JALCIVAR	05/01/2025 21:09:38 GMT
		Reason Code Test				JOHANN ALCIVAR	
450.00	Maximum Dose	18.2 kGy	40.0 kGy	34.4 KGY	Pass	JALCIVAR	05/01/2025 21:09:56 GMT
		Reason Code Test				JOHANN ALCIVAR	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: BRITTANY LONG  
 Reason: Work Order Completions

Date: 05/02/2025 13:03:18 GMT