



CERTIFICATE OF CONFORMANCE AND STERILITY

This certification is provided as full assurance that the following product code and lot number were manufactured in accordance via a Quality System certified to ISO 9001 with prescribed procedures and specifications.

Product Code:	MEQT0002SZ
Lot Number:	022846
Manufacturing Date:	2024.12.27
Expiration Date:	2026.12
Sterility Assurance Level:	10 ⁻⁶ (SAL determined by test method ISO11137)
Radiation Lot:	25010704
Irradiation Dose Specified:	25 kGy to 50 kGy
Minimum Absorbed Dose:	28.9 kGy
Maximum Absorbed Dose:	35.7 kGy
Radiation Date:	2025.01.07

Quality Department
Contec Cleanroom Technology (Suzhou) Co., Ltd.
2025-01-09



NORTH AMERICA

Contec, Inc.
525 Locust Grove
Spartanburg, SC 29303 USA
tel: +1 864 503 8333
fax: +1 864 503 8444

EUROPE

Contec Europe
ZI Du Prat, Avenue Paul Duplex
56000 VANNES France
tel: +33 (0)2 97 43 76 98
fax: +33 (0) 97 43 76 86

ASIA

Contec Cleanroom Technology (Suzhou) Co., Ltd.
17 Longyun Road
Suzhou Industrial Park, Suzhou 215024 China
tel: +86 512 6274 4050
fax: +86 512 6274 4051

web: www.contecinc.com
e-mail: info@contecinc.com

证书编号:

 Certificate No.: **0040968**

辐照证明书

CERTIFICATE OF IRRADIATION

执照登记号(Irradiation License No.):91320582743739332H

顾客:

康泰无尘科技(苏州)有限公司

编号: CHK/QR-8.2.6-02

Customer:

版本/D/1

产品名称 Product	包装规格(厘米) Dimension(cm)	产品批号 Product Lot No.	数量(箱) Qty.(ctns)	毛重(公斤) G.W. (kg)
无尘拖把套MEQT0002SZ	----	022846	1607 箱	----
无尘包装袋BKT BAG2	----	022797	250 箱	----
无尘包装袋BKT BAG	----	022796	200 箱	----
无尘拖把套MEQT1002	----	022763	10 箱	----
辐照日期(Irradiation start 20250107)				
结束日期(Irradiation finished 20250108)				
以下为空白(The following is blank)				

张家港市中核华康辐照有限公司兹证明以上所列的所有产品(同制造商描述)在放射量测定系统精度和准确度的使用范围内接受到如下剂量。Zhangjiagang Municipal CNNC Huakang Radiation Co.,Ltd certifies that the products listed above described by the manufacturer receiver the indicated doses within the precision and accuracy of the dosimetry system employed 剂量计测量系统的总不确定度为 5% The uncertainty of the dosimeter measurement system is 5%

辐照批号 Irradiation Lot No.	25010704	辐照完成日期 Irradiation finished at	20250108
顾客要求最低吸收量 Specified minimum dose	25 kGy	实际最低吸收剂量 Minimum absorbed dose	28.9 kGy
顾客要求最高吸收量 Specified maximum dose	50 kGy	实际最高吸收剂量 Maximum absorbed dose	35.7 kGy
剂量计型号及批号 Dosimeter type & batch	重铬酸钾(银) 240910	张家港市中核华康辐照有限公司 Zhangjiagang Municipal CNNC Huakang Radiation Co.,LTD 地址(ADD): 张家港市凤凰镇西张创业路28号 No.28,Chuangye Road,Fenghuang Town,Zhangjiagang City 电话(Tel) 0512-58421200 传真(Fax) 0512-58455118 邮编(P.C.) 215614	
检验员/日期: Inspected by:	肖卓凌 20250108		
批准人/日期: Approved by:	沈少明 20250108		

本证明书签字盖章有效

The certificate is valid after signature and stamped.

NORTH AMERICA

 Contec, Inc.
 525 Locust Grove
 Spartanburg, SC 29303 USA
 tel: +1 864 503 8333
 fax: +1 864 503 8444

EUROPE

 Contec Europe
 ZI Du Prat, Avenue Paul Duplex
 56000 VANNES France
 tel: +33 (0)2 97 43 76 98
 fax: +33 (0) 97 43 76 86

ASIA

 Contec Cleanroom Technology (Suzhou) Co., Ltd.
 17 Longyun Road
 Suzhou Industrial Park, Suzhou 215024 China
 tel: +86 512 6274 4050
 fax: +86 512 6274 4051

 web: www.contecinc.com
 e-mail: info@contecinc.com