



Certificate of Conformance, Analysis, and Sterility

This certification is provided as full assurance that the following product code and lot number were manufactured in accordance via a Quality System certified to ISO 9001 with prescribed procedures and specifications.

Product Code:	PSC20001-BPR
Lot Number:	1102560-0-0-1
EU BPR Authorisation Number:	EU-0020460 0002 (1-2)
Manufacturing Date:	01/08/2025
Expiration Date:	01/2027

Analysis

% v/v Isopropyl Alcohol:	69.989
--------------------------	--------

Sterility

Gamma Processing Run ID:	3689811
Irradiation Dose Specified:	25.0 kGy to 50.0 kGy
Irradiation Dose Delivered:	28.1 kGy to 37.2 kGy
Validated Sterile Product:	Validation Sterile packets available upon request.
Sterility Assurance Level:	10 ⁻⁶ (SAL determined by test method AAMI/ISO 11137)

QC Supervisor
coc@contecinc.com

Tuesday, January 21, 2025

NORTH AMERICA

Contec, Inc.
525 Locust Grove
Spartanburg, SC 29303 USA
tel: +1 864 503 8333
fax: +1 864 503 8444

EUROPE

Contec Europe
ZI Du Prat, Avenue Paul Dupleix
56000 VANNES France
tel: +33 (0)2 97 43 76 98
fax: +33 (0) 97 43 76 86

ASIA

Contec Cleanroom Technology (Suzhou) Co., Ltd.
17 Longyun Road
Suzhou Industrial Park, Suzhou 215024 China
tel: +86 512 6274 4050
fax: +86 512 6274 4051

web: www.contecinc.com
e-mail: info@contecinc.com

Page 1 of 1

DID: 20605A



Certificate of Processing

STERIGENICS 10811 Withers Cove Park Dr. Charlotte NC 28278
 TEL 704 588-6877 FAX 704 588-8355 www.sterigenics.com

R55480102

01/20/2025 15:49:02 GMT
 Page 1 of 1

Customer Name:	Contec- PO Box 530	Processing Facility:	Charlotte Gamma	Work Order #	3689811
P.O.#	146648			Sales Order #	3720098
25-50 kGy		CNT1, Gamma Treatment	Irradiation Date/Time:	01/17/2025 16:51:00 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	108	CA	PSC20001-BPR	Clean Room Product	1102560-0-0-1	146648
102.000	108	CA	SB327030IR	32 oz IPA Spray Bottle 70/30	1096463-0-0-1	146648
	216	CA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By ----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	50.0 kGy	28.1 KGY	Pass	BPOSTON	01/18/2025 02:33:36 GMT
		Reason Code Test				BRANTT POSTON	
450.00	Maximum Dose	25.0 kGy	50.0 kGy	37.2 KGY	Pass	BPOSTON	01/18/2025 02:34:00 GMT
		Reason Code Test				BRANTT POSTON	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: WENDY KOBLISH
 Reason: Work Order Completions

Date: 01/20/2025 15:21:07 GMT