

Certificate of Conformance, Analysis, and Sterility

This certification is provided as full assurance that the following product code and lot number were manufactured in accordance via a Quality System certified to ISO 9001 with prescribed procedures and specifications.

Product Code:	PS-911EB-BPR
Lot Number:	1128469-0-0-1
Description:	PROSAT sterile pre saturated wipes
EU BPR Authorisation Number:	EU-0020460 0002 (1-2)
Manufacturing Date:	02/28/2025
Isopropyl Alcohol:	70 ± 3% by volume with the remainder DI H2O.
Expiration Date:	02/2027

Analysis

Quantity of Particles Po 0.5µM:	3.732 millions/m ²
---------------------------------	-------------------------------

Sterility

Gamma Processing Run ID:	3721849
Irradiation Dose Specified:	18.2 kGy to 40.0 kGy
Irradiation Dose Delivered:	20.0 kGy to 31.5 kGy
Validated Sterile Product:	Validation Sterile packets available upon request.
Sterility Assurance Level:	10 ⁻⁶ (SAL determined by test method AAMI/ISO 11137)



QC Supervisor
coc@contecinc.com

Monday, March 17, 2025

NORTH AMERICA

Contec, Inc.
525 Locust Grove
Spartanburg, SC 29303 USA
tel: +1 864 503 8333
fax: +1 864 503 8444

EUROPE

Contec Europe
ZI Du Prat, Avenue Paul Duplex
56000 VANNES France
tel: +33 (0)2 97 43 76 98
fax: +33 (0) 97 43 76 86

ASIA

Contec Cleanroom Technology (Suzhou) Co., Ltd.
17 Longyun Road
Suzhou Industrial Park, Suzhou 215024 China
tel: +86 512 6274 4050
fax: +86 512 6274 4051

web: www.contecinc.com
e-mail: info@contecinc.com

Page 1 of 1

DID: 20602A



Certificate of Processing

STERIGENICS 10811 Withers Cove Park Dr. Charlotte NC 28278
TEL 704 588-6877 FAX 704 588-8355 www.sterigenics.com

R55480102

03/13/2025 13:33:07 GMT
Page 1 of 1

Customer Name: Contec- PO Box 530
P.O.# 148146

Processing Facility: Charlotte Gamma

Work Order # 3721849
Sales Order # 3757668

18.2-40.0 kGy

CNT4, Gamma Treatment

Irradiation Date/Time:

03/12/2025 06:10:00 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	414	CA	PS-911EB-BPR	PROSAT Sterile Wipes	1128469-0-0-1	148146
102.000	432	CA	HCPS2328IR	Polyester Wipes	1107824-0-0-1	148146
	846	CA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	18.2 kGy	40.0 kGy	20.0 KGY	Pass	JALCIVAR	03/12/2025 15:18:43 GMT
		Reason Code Test				JOHANN ALCIVAR	
450.00	Maximum Dose	18.2 kGy	40.0 kGy	31.5 KGY	Pass	JALCIVAR	03/12/2025 15:19:00 GMT
		Reason Code Test				JOHANN ALCIVAR	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: BRITTANY LONG
Reason: Work Order Completions

Date: 03/13/2025 13:30:57 GMT