
Certificate of Conformity

Product: CryoVial Cryo Color Codes, Internal thread, Green

Date: 2017-05-04

Page: 1

Item number: 377251

Lot number: 152112

Date of manufacture: April 2017

Date of expiry: April 2022

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2008 "Quality management system - Requirements" and ISO 13485:2003 "Medical devices - Quality management systems - System requirements for regulatory purposes".

Roskilde Site



Anne-Mette Reimann
QA Manager