

Catalog Number: 01549750

Lot Number: 601215

Product Description: Fisherbrand™ Sterile Standard Weighing Boat

Irradiation lot No.: 00005961

Irradiation Finished Date: 07/01/2024

Dose Inspection Result: 25.02 kGy

Specified Does: 25-50kGy

Expiration Date: 07/31/2027

Certificate of Irradiation

Products have been irradiated with high-energy electron beams.

Our supplier is committed to developing and manufacturing high-quality, industry-leading products for scientific research. We certify that the following product has been produced and inspected according to our standard quality assurance processes.

Their mission is to provide conscientiously designed and thoroughly tested products for laboratory professionals in the pure and applied sciences. They strive to maintain the highest quality standards and welcome all inquiries in this regard.

This Certificate has been compiled with data that has been provided by the manufacturer of the product. Fisher Scientific Channel has not independently verified such data. This document has been produced electronically and is valid without a signature.