

# CERTIFICATE OF ANALYSIS

## ***Standardized Solution***

**Description:** Sulphuric Acid 0.05M (0.1N)

**Lot N:** 1123181

**Ref N:** 12903634

**Measured Concentration:** **0.04996 mol/l +/- 0.2%**

**Traceability:** BAM 93440 Lot BCCK0309

**Method:** WQP 5.15.1/13

**Valid until:** 05.07.2027

This solution was produced using quality Ph Eur components and water meeting the Ph Eur monograph requirement for Purified Water.

*Instruction for use*

The solution bottle should be open for the minimum time required to dispense the solution. After use, the bottle should be tightly recapped and stored under normal laboratory conditions.

This certificate relates solely to the lot number given above.

**Certification date:** 05.06.2025

*Signed by: *, Chemical Production Manager