

CERTIFICATE OF ANALYSIS

Standardized Solution

Description: Sulphuric Acid 0.05M (0.1N)

Lot N: 1150322

Ref N: 12903634

Measured Concentration: 0.04996 mol/l +/- 0.2%

Traceability: BAM 93440 Lot BCKK0309

Method: WQP 5.15.1/13

Valid until: 01.10.2027

This solution was produced using quality Ph Eur components and water meeting the Ph Eur monograph requirement for Purified Water.

Instruction for use

The solution bottle should be open for the minimum time required to dispense the solution. After use, the bottle should be tightly recapped and stored under normal laboratory conditions.

This certificate relates solely to the lot number given above.

Certification date: 01.09.2025

Signed by: , Chemical Production Manager