

CERTIFICATE OF ANALYSIS

European Pharmacopoeia Phenol red solution

Lot N:	1103186
Description:	Ph Eur N 1063601
Ref N:	12930484
Starting materials:	phenol red R / 0.1 M sodium hydroxide / ethanol (96 per cent) R / water R.
Test for sensitivity:	passed
Colour change:	pH 6.8 (yellow) to pH 8.4 (reddish-violet)
Expiry date:	02.05.2027

The reagent is specially intended for use in testing Ph Eur and monographs. This solution was produced according 4.1.1 of Ph Eur specifications using Ph Eur reagent quality components and water meeting the Ph Eur monograph requirement for Purified Water.

Instruction for use

The solution bottle should be open for the minimum time required to dispense the solution. After use, the bottle should be tightly recapped and stored under ambient temp.

This certificate relates solely to the lot number given above.

Certification date:	02.04.2025
----------------------------	------------

Signed by:  , Chemical Production Manager