

# Certificate of Analysis

## *European Pharmacopoeia Purified water in bulk*

**Lot N:** 2506015

**Description:** Ph Eur N 1095500

**Ref N:** 12940564

**Performance test:** Microbiological monitoring - passed

Total organic carbon or oxidisable substances - passed

Conductivity - 0.8 uS/cm at 25°C

Nitrates - < 0.2ppm

Aluminium - < 10ppb

Heavy metals - < 0.1 ppm

**Expiry date:** 31.07.2026

The reagent is specially intended for use in testing Ph Eur and monographs. This solution is not suitable for use in the manufacture of dialysis solutions. It is not sterile and apyrogenic.

### *Instructions for use*

The bottle should be open for the minimum time required to dispense the solution. After use, the bottle should be tightly recapped and stored under normal laboratory conditions.

This certificate relates solely to the lot number above.

**Certification date:** 10.06.2025

Signed by: . . . . . *G. Mann* . . . . ., Chemical Production Manager