

CERTIFICATE OF ANALYSIS

CERTIFIED REFERENCE MATERIAL

7 components: 10mg/l each of Bromides (Br-) ; Chlorides (Cl-) ; Fluorides (F-) ; Nitrates (NO₃-) ; Nitrites (NO₂-) ; Phosphates (PO₄3-) ; Sulphates (SO₄2-) in Water

Lot N: 1129926
Barcode: 17033770

Ref N: 12942235

Certification Date: 26.06.2025

Component	Certified Value and uncertainty [mg/l]	Metrological traceability
F ⁻	10.031 ± 0.032 ^(y)	NIST SRM No 3183 Lot 140203
Cl ⁻	9.972 ± 0.024 ^(y)	NIST SRM No 3182 Lot 190830
NO ₂ ⁻	9.991 ± 0.036 ^(y)	BAM CRM No 71804 Lot BCCD7688
Br ⁻	9.986 ± 0.027 ^(y)	NIST SRM No 3184 Lot 151130
NO ₃ ⁻	10.005 ± 0.024 ^(y)	NIST SRM No 3185 Lot 170309
SO ₄ ²⁻	9.982 ± 0.023 ^(y)	NIST SRM No 3181 Lot 080603
PO ₄ ³⁻	10.051 ± 0.033 ^(y)	NIST SRM No 3186 Lot 170606

Notes:

(y) WQP 5.15.1.24 The certified value was obtained by a weighted mean of the results of two independent testing methods among: Classical Volumetric, Primary Gravimetric, Instrumental (ICP, ICP/MS or IC)

Density* 0.997 g/cm³ at 20°C

Starting Material, Purity*	Batch
KBr 99.999%	82118822
NaCl 99.999%	82089719
NaF 99.995%	82118242
NaNO ₃ 99.999%	82118693
NaNO ₂ 99.994%	82107833
NaH ₂ PO ₄ 99.993%	82089801
Na ₂ SO ₄ 99.999%	82085834

* These values are not certified

Storage Conditions: Store under normal laboratory conditions, at temperatures between 15°C to 25°C

Expiry date: 26.07.2027

Concept of Certification and traceability statement:

This certified reference material (CRM) is produced using a high purity starting material, acid from sub-boiling and 18 MOhm deionised water and filtered through a 0.2 micron filter.

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with EA 4/02

Property of the result of a measurement whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties (ISO VIM)

The metrological traceability is assured using certified reference material traceable to SI of NIST (SRM) or BAM (CRM). All contributions in relation to the certification of standard solutions are considered when evaluating the uncertainty.

The measurement results are traceable to SI.

All analytical balances used for the preparation of the solution are calibrated yearly under an in-house procedure with analytical weights, traceable to DKD, and are checked daily.

Class A laboratory glassware is used.

The results from temperature measurement are traceable to SI. The thermometers used for solution's calibration are calibrated from an ISO 17025 accredited laboratory. The ambient conditions are controlled with a hygrometer calibrated from an ISO 17025 accredited laboratory.

Intended use: For Laboratory Use Only

Calibration of Ion Chromatographs
Preparation of "working reference samples"

Validation of analytical methods
Detection limit and linearity studies

This statement is not intended to restrict the use for other purposes.

Instructions for the correct use of this reference material:

This certified reference material can be used directly or can be diluted in an appropriate high purity matrix. Only a clean class A glassware should be used. Do not pipet from container. Obtained concentration (in mg/l) after dilution is a result from the multiplication of certified value of CRM concentration and the CRM's volume used for dilution and divided into the flask's volume used for dilution.

Stability and storage:

This CRM is with a guaranteed stability until $\pm 0.5\%$ of the certified concentration within its shelf life. Stability is guaranteed, provided that the solution is kept in its original packaging, tightly closed stored, as written in the section: Storage Conditions. If storage of a partially used bottle is necessary, the cap should be tightly sealed and the bottle should be stored in refrigerator to minimize transpiration rate. The laboratory performs stability tests according to MQP 5.14.1 therefore solutions with one and the same bar-code number might have different expiration dates.

Hazardous situation:

The normal laboratory safety precautions should be observed when working with this CRM. Further details for the handling of this CRM are available as safety data sheet.

Level of homogeneity:

This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous.
To ensure sufficient homogeneity of the sample prior to use thoroughly mix by inversion.

Names of certifying officers:

Laboratory:



Tihomir Stoyanov

Manager:



Krassimira Taralova

This document QF 5.17.1/1 version 1 is designed and the certified value(s) and uncertainty(ies) are determined in accordance with ISO Guide 31, ISO Guide 35, and Eurachem / CITAC Guides

This certificate relates solely to the lot number given above.

All processes (including generating of this certificate) are completely controlled by the specialized Computer-Aided-Manufacturing (CAM) software.

This Certified Reference Material was produced under a quality management system that is:

- Registered to ISO 9001 Quality Management System (Lloyd's Register Quality Assurance Ltd Cert No 0039638)
- Accredited according to ISO/IEC 17025
- Accredited according to ISO 17034