

CERTIFICATE OF ANALYSIS

BioProcess Container Systems

Product: 10L LABTAINER W/QUICK CONNECTS, CX5-14

Lot #: BLF117245
Drawing #: 06266

Catalog #: SH30712.02
Drawing Rev: "J"

Expiration Date: JUN/2028
Manufacture Date: JUN/2025

Test	Specification	Units	Results
Irradiation Dose	25.0 – 45.0	kGy	Pass

- ✦ **Inspection:** This lot of BioProcess Container Systems has been 100% visually inspected per specification.
- ✦ **Biological Reactivity:** All product contact materials have passed USP Class VI testing (USP <88>) and/or ISO10993.
- ✦ **Cytotoxicity:** All Thermo Scientific product contact films have passed cytotoxicity testing (USP <87> MEM Elution)
- ✦ **Physicochemical:** All Thermo Scientific product contact films have passed USP Physicochemical Tests for Plastics (USP <661>).
- ✦ **EP Testing:** All Thermo Scientific product contact films have passed EP <3.2.2.1> "Plastic Containers for Aqueous Solutions for Parenteral Infusion".
- ✦ **Endotoxin:** Samples of representative BioProcess Container Systems are routinely tested in periodic validations for the presence of endotoxin per the USP Bacterial Endotoxin Test (USP <85>). Aqueous extracts contained <0.25 EU/ml as determined by the Limulus Amebocyte Lysate Test (LAL).
- ✦ **Particulate:** Samples of representative BioProcess Container Systems have been routinely tested in periodic validations and have passed requirements per Particulate Matter in Injections Light Obscuration Particulate Count Test (USP <788>).
- ✦ **Sterility:** Routine sterility testing is performed on representative samples of Thermo Scientific BioProcess Container Systems following ANSI/AAMI/ISO 11137 guidelines. Periodic validation has determined that an irradiation dose of 25.0 – 45.0 kGy provides a minimum Sterility Assurance Level (SAL) of 10^{-6} for product contact surfaces.

Rinda Potts 16 Jul 2025

Quality Department / Date Issue

Certificate of Processing

STERIGENICS 1700 College Boulevard West Memphis AR 72301
 TEL 870 735-8400 FAX 870 735-9977 www.sterigenics.com

R55480102

07/09/2025 18:47:24 GMT
 Page 1 of 1

Customer Name: Thermo Fisher (Life Tech.)- 1726 HyClone Processing Facility: West Memphis Work Order # 3790281
 P.O.# 278889 Sales Order # 3841992
 25.0 - 45.0 kGy THF-A, Gamma Treatment Irradiation Date/Time: 07/07/2025 22:17:03 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	20	CA	18X14X6	500 mL Labtainer	BLF124195	LEB_SH30658.15
	Dose Map		138_2222	MIN 0C1 MAX 16E7		
102.000	10	CA	M 36x23x12	Labtainer	BLF117248	LEB_SH30967.02
	Dose Map		138_2224	REF(TC1) AFmin0.822 AFmax1.123		
103.000	21	CA	M 36x23x12	Labtainer	BLF117247	LEB_SH30967.02
	Dose Map		138_2224	REF(TC1) AFmin0.822 AFmax1.123		
104.000	1	CA	M 36x23x12	Labtainer	BLF633401	LEB_SUT15333
	Dose Map		138_2224	REF(TC1) AFmin0.822 AFmax1.123		
105.000	50	CA	M 36x23x12	Labtainer	BLF117245	LEB_SH30712.02
	Dose Map		138_2224	REF(TC1) AFmin0.822 AFmax1.123		
106.000	60	CA	L 47x23x12	20x20 Labtainer	BLF118980	LEB_SH3B1323.04
	Dose Map		138_2230	MIN-1C1 MAX-12E7		
	162	CA	Total			

Processing Summary

Op#	Operation Name	Location	Date/Time In (GMT)	Date/Time Out (GMT)
400.00	Receive Product	UNPROC	07/01/2025 13:15:00	07/01/2025 17:23:46
410.00	Prepare Product	UNPROC	07/01/2025 17:23:46	07/07/2025 22:17:03
420.00	Load Product	UNPROC	07/07/2025 22:17:03	07/08/2025 05:30:21
430.00	Process Product	UNPROC	07/08/2025 05:30:21	07/08/2025 05:30:26
440.00	Unload Product	UNPROC	07/08/2025 05:30:26	07/08/2025 05:30:30
450.00	Collect Run Data	UNPROC	07/08/2025 05:30:30	07/08/2025 23:37:49

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	45.0 kGy	27.6 KGY	Pass	JCLAY	07/09/2025 17:03:24 GMT
	Reason Code Test					JAMES D CLAY	
450.00	Maximum Dose	25.0 kGy	45.0 kGy	37.7 KGY	Pass	JCLAY	07/09/2025 17:03:24 GMT
	Reason Code Test					JAMES D CLAY	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Handwritten signature: J.C.P. 16 JUL 2025

Electronically Signed By: DEBRA MACK
 Reason: Work Order Completions

Date: 07/09/2025 18:46:56 GMT