

## Certificate of Conformity

**Product:** Nunclon Sphera 24 Well Multidish  
**Item Number:** 174930  
**Lot number:** 132404  
**Date of manufacture:** September 2024  
**Date of expiry:** September 2028

Date: 2024-11-08  
Page: 1

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies
USP<85> Bacterial endotoxins test	<0.5 endotoxins unit/ml	Complies
USP <88> "Biological Reactivity test In Vivo, USP Class VI Test"	USP current edition	Complies**
Protein adsorption	Max 2.0 %	Complies
Cell adhesion	Max 1.0 %	Complies

\*\*Only performed once

Sterility is obtained by ethylene oxide gas according to ISO 11135-1 "Sterilization of health care products - Ethylene oxide", with SAL 10E-6.

Roskilde Site



Mette Mikkelsen

Quality Control Coordinator

Please note that this is an electronically generated signature

Certificate reviewed by  
Roskilde Site



Sabrina Hammeken Nissen

QC Technician II

Please note that this is an electronically generated signature