

---

## Certificate of Irradiation

<b>Product</b>	Nunc F96 Microwell plate w/lid, clear	<b>Date</b> 2025-03-07
		<b>Page</b> 1
<b>Item Number</b>	260860	
<b>Lot Number</b>	185137	
<b>Specified Dose Range</b>	16.4 - 29.4 kGy	
<b>SAL</b>	10E-3	
<b>Quantity</b>	180 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EA__1719892A	DK00517061	2025-02-28	18.9
EA__1719896A	DK00517063	2025-02-28	19.3
EA__1720104A	DK00517062	2025-02-28	19.3
EA__1720310A	DK00517064	2025-02-28	18.7
EA__1720309A	DK00517065	2025-02-28	18.9
EA__1719945B	DK00517066	2025-02-28	19.1
EA__1722485B	DK00517067	2025-03-06	19.9
EA__1722279B	DK00517069	2025-03-06	19.2
EA__1722260B	DK00517068	2025-03-06	19.4

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack  
Senior Quality Assurance Manager

Please note that this is an electronically generated signature