
Certificate of Irradiation

Product	Nunc Dish Nunclon 35 x 10	Date 2025-05-18
		Page 1
Item Number	150318	
Lot Number	185815	
Specified Dose Range	15.8 - 44.3 kGy	
SAL	10E-3	
Quantity	400 boxes	

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications.

Irradiation lot no.	PalletID	Date of irradiation	Dose kGy
EC__1300084A	DK00397567	2025-05-17	18.7
EC__1300098A	DK00397566	2025-05-17	18.1
EC__1300137B	DK00397565	2025-05-17	18.2
EC__1300028A	DK00397564	2025-05-17	18.5
EC__1300056A	DK00397563	2025-05-17	18.9

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation".

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality Management System - Requirements" and ISO 13485:2016 "Medical Devices - Quality Management Systems - System requirements for regulatory purposes".

Roskilde Site



Per Bernack
Senior Quality Assurance Manager

Please note that this is an electronically generated signature