

**Visual Inspection
– Barrier Defects
(Holes/Pin Holes)**

Inspection Level	S4	AQL	1.5	Maximum Allowable	5
------------------	----	-----	-----	-------------------	---

SIZE	SAMPLE SIZE	CUFF	PALM	CROTCH	FINGER	F/TIP	OTHERS	TOTAL
XS	0							0
S	0							0
M	125	1	0	0	0	0	0	1
L	0							0
XL	0							0
XXL	0							0
Total	125	1	0	0	0	0	0	1

ACC	
REJ	

**Liquid Particle Counts
(counts/cm²)**

Equipment: Using Liquid Syringe Sampler, Brand: PMS, Model # SLS 1500, Test Method IES-RP-CC005.4

SIZE	0.5-1.0um	1.0-2.0um	2.0-5.0um	5.0-10um	10-20um	< 20um	Total	Spec counts/cm ²
XS	0	0	0	0	0	0	0	< 2200
S	0	0	0	0	0	0	0	< 2200
M	1215	92	20	7	3	0	1337	<2200
L	0	0	0	0	0	0	0	<2200
XL	0	0	0	0	0	0	0	<2200
XXL	0	0	0	0	0	0	0	<2200

**Ion Chromatograph Determination
(ug/cm²)**

Equipment: Ion Chromatograph, Brand: Tosoh, Model # IC-2010, Test Method IES- RP-CC005.4

ANION (ug/cm ²)	SIZE						Spec
	XS	S	M	L	XL	XXL	
Fluoride	0	0	<0.0004	0	0	0	<0.0400
Chloride	0	0	0.1081	0	0	0	<1.4000
Nitrate	0	0	0.1369	0	0	0	<1.2000
Sulphate	0	0	0.0132	0	0	0	<1.0000
CATION (ug/cm ²)	SIZE						Spec
	XS	S	M	L	XL	XXL	
Sodium	0	0	0.0104	0	0	0	<0.0600
Magnesium	0	0	0.0017	0	0	0	<0.0050
Lithium	0	0	<0.0003	0	0	0	<0.0005
Copper	0	0	<0.0003	0	0	0	<0.0005
Potassium	0	0	0.0159	0	0	0	<0.0600

Non-Volatile Residue Determination	<p><i>Equipment: Using Mettler Toledo instrument model UMX2 and using IPA, Test Method IES-RP-CC005.4</i></p> <table border="1" data-bbox="410 195 1295 331"> <thead> <tr> <th rowspan="2">Parameter (ug/cm²)</th> <th colspan="6">SIZE</th> <th rowspan="2">Spec</th> </tr> <tr> <th>XS</th> <th>S</th> <th>M</th> <th>L</th> <th>XL</th> <th>XXL</th> </tr> </thead> <tbody> <tr> <td>NVR</td> <td>-</td> <td>-</td> <td>1.07</td> <td>-</td> <td>-</td> <td>-</td> <td>< 3.0</td> </tr> </tbody> </table>	Parameter (ug/cm ²)	SIZE						Spec	XS	S	M	L	XL	XXL	NVR	-	-	1.07	-	-	-	< 3.0																
Parameter (ug/cm ²)	SIZE						Spec																																
	XS	S	M	L	XL	XXL																																	
NVR	-	-	1.07	-	-	-	< 3.0																																
FTIR Determination	<p><i>Equipment: FTIR instrument model FTIR-8400, Test Method IES-RP-CC005.4</i></p> <table border="1" data-bbox="410 436 1393 646"> <thead> <tr> <th rowspan="2">Parameter (ug/cm²)</th> <th colspan="6">SIZE</th> <th rowspan="2">Spec</th> </tr> <tr> <th>XS</th> <th>S</th> <th>M</th> <th>L</th> <th>XL</th> <th>XXL</th> </tr> </thead> <tbody> <tr> <td>Amide</td> <td>-</td> <td>-</td> <td>Negative</td> <td>-</td> <td>-</td> <td>-</td> <td>Negative</td> </tr> <tr> <td>DOP</td> <td>-</td> <td>-</td> <td>Negative</td> <td>-</td> <td>-</td> <td>-</td> <td>Negative</td> </tr> <tr> <td>Silicone Oil</td> <td>-</td> <td>-</td> <td>Negative</td> <td>-</td> <td>-</td> <td>-</td> <td>Negative</td> </tr> </tbody> </table>	Parameter (ug/cm ²)	SIZE						Spec	XS	S	M	L	XL	XXL	Amide	-	-	Negative	-	-	-	Negative	DOP	-	-	Negative	-	-	-	Negative	Silicone Oil	-	-	Negative	-	-	-	Negative
Parameter (ug/cm ²)	SIZE						Spec																																
	XS	S	M	L	XL	XXL																																	
Amide	-	-	Negative	-	-	-	Negative																																
DOP	-	-	Negative	-	-	-	Negative																																
Silicone Oil	-	-	Negative	-	-	-	Negative																																
Status	<table border="1" data-bbox="410 751 1393 961"> <tbody> <tr> <td>VISUAL DEFECTS – MAJOR & MINOR</td> <td>PASS</td> </tr> <tr> <td>VISUAL DEFECTS – BARRIER</td> <td>PASS</td> </tr> <tr> <td>LIQUID PARTICLE COUNTS</td> <td>PASS</td> </tr> <tr> <td>ION CHROMATOGRAPH</td> <td>PASS</td> </tr> <tr> <td>NON – VOLATILE RESIDUE</td> <td>PASS</td> </tr> <tr> <td>FTIR</td> <td>PASS</td> </tr> </tbody> </table>	VISUAL DEFECTS – MAJOR & MINOR	PASS	VISUAL DEFECTS – BARRIER	PASS	LIQUID PARTICLE COUNTS	PASS	ION CHROMATOGRAPH	PASS	NON – VOLATILE RESIDUE	PASS	FTIR	PASS																										
VISUAL DEFECTS – MAJOR & MINOR	PASS																																						
VISUAL DEFECTS – BARRIER	PASS																																						
LIQUID PARTICLE COUNTS	PASS																																						
ION CHROMATOGRAPH	PASS																																						
NON – VOLATILE RESIDUE	PASS																																						
FTIR	PASS																																						
Conclusion	<p>Based on the results reported herein, obtained from samples selected randomly accordingly to ISO 2895, the lot inspected having fulfilled all the product specifications, is hereby certified to have complied with the desired product requirements.</p>																																						

This Certificate has been compiled with data that has been provided by the manufacturer of the product. The Fisher Scientific channel has not independently verified such data. This document has been produced electronically and is valid without a signature.

Fisher Scientific Company L.L.C. Quality Assurance Department