

Catalog Number: 02707400

Lot Number: 23140262

Description: Fisherbrand™ SureOne™ Micropoint Pipette Tips, Universal Fit, Non-Filtered

Certificate of Conformity

Manufacturing Date: 04/12/2023

Irradiation Date: N/A

Sterility Assurance Level (SAL):N/A

Sterility Assurance Level for sterilized products is ensured by irradiating the product according to ANSI/AAMI/ISO 11137- 2015 guidelines.

Our supplier, hereby certifies that these are manufactured in a controlled environment in accordance with established products manufacturing guidelines, quality requirements and product specifications. They are certified, by validation of the manufacturing processes, to be free of the following contaminants:

- Human and mouse DNA (Deoxyribonucleic Acid) : < 30 pg
- Ribonuclease (RNase): < 10 Kunitz Units/ul
- Deoxyribonuclease (DNase): < 1077 Kunitz Units/l
- PCR Inhibitors: not present
- Endotoxin (Pyrogen) < 0.06 EU/ml
- Adenosine-5'-triphosphate (ATP) < 107 mg/ul

This Certificate has been compiled with data that has been provided by the manufacturer of the product. The Fisher Scientific channel has not independently verified such data. This document has been produced electronically and is valid without a signature.