Certificate of Conformity

Lot Number: 24500275

Part Number: 509-GRD-R-PFB

Item Description: 1.5ml FIS PREM MCT 05-408-1301.5ml MCT Graduated - Red

Manufacturing Date: 12/18/2024

Irradiation Date: N/A

Sterility Assurance Level (SAL):N/A

Sterility Assurance Level for sterilized products is ensured by irradiating the product according to ANSI/AAMI/ISO 11137- 2015 guidelines.

Molecular BioProducts Inc., a part of Thermo Fisher Scientific Inc., hereby certifies that these products are manufactured in a controlled environment in accordance with established manufacturing guidelines, quality requirements and product specifications. They are certified, by validation of the manufacturing processes, to be free of the following contaminates:

- Human and mouse DNA (Deoxyribonucleic Acid) : < 30 pg
- Ribonuclease (RNase): < 10⁻⁹ Kunitz Units/µl
- Deoxyribonuclease (DNase): < 10⁻⁷ Kunitz Units/µl
- PCR Inhibitors: not present
- Endotoxin (Pyrogen) < 0.06 EU/ml
- Adenosine-5'-triphosphate (ATP) < 10⁻¹³ mg/µl

Q. Ibarria

Certified by:

12/27/2024

Ana Ibarria Interim Quality Manager

Laboratory Products Division Document No.: FRM0033085 Rev J.00 9389 Waples Street Area: Quality San Diego, CA(800) 995-278792121-3903(858) 453-4367 fax

www.thermofisher.com ISO 9001:2015 ISO13485:2016 FRM0033085 REV. J.00 Ref. WIN0021995