

CERTIFICATE OF ANALYSIS

| | |
|-----------------------------------|----------------|
| PRODUCT NAME | Isofield LS300 |
| PRODUCT CODE / SIZE | 10230 M |
| LOT | 240291 |
| MANUFACTURING DATE | 2024-06 |
| EXPIRY DATE | 2029-05 |
| CERTIFICATE OF IRRADIATION | NA |

| PROPERTY | TEST METHOD | TEST PLAN | RESULT |
|-----------------------------|------------------------|---------------|----------|
| Pinholes – Water Tight Test | EN ISO 374-2 | AQL 1.5; G1 | Conforms |
| Visual – major defects | In-house | AQL 2.5; G1 | Conforms |
| Visual – minor defects | In-house | AQL 4.0; G1 | Conforms |
| Dimension | EN ISO 21420, EN 455-2 | AQL 4.0; S2 | Conforms |
| Physical Properties | EN455-2 | AQL 4.0; S2 | Conforms |
| Packing Configuration | In-house | NA | Conforms |
| Extractable Ionics | IENT-RP-CC005.4 | N=3 | Conforms |
| Liquid Particle Count | IENT-RP-CC005.4 | N=3 | Conforms |
| Sterilization | Gamma Irradiation | Minimum 25kGy | NA |
| Endotoxin | EN455-3, USP | ≤20 EU/Pair | Conforms |



COMPANY

ISOFIELD SDN. BHD.
(Registration No: 201601039008 (1348338-A))

Authorized Signatory

NAME

Steve Morgan

DATE

12 June 24

CERTIFICATE OF CONFORMANCE

| | |
|----------------------------|----------------|
| PRODUCT NAME | Isofield LS300 |
| PRODUCT CODE / SIZE | 10230 M |
| LOT | 240291 |
| MANUFACTURING DATE | 2024-06 |
| EXPIRY DATE | 2029-05 |
| CERTIFICATE OF IRRADIATION | NA |



COMPANY

ISOFIELD SDN. BHD.
(Registration No: 201601039008 (1348338-A))

.....
Authorized Signatory

NAME

Steve Morgan

DATE

12 June 24

The above product is hereby declared to have been inspected and found to be in conformance with the agreed specifications as reported in the Certificate of Analysis.