

<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 05-Jul-2021

MY03S12583214-1-1

This is to certify that Synergy Sterilisation (M) Sdn. Bhd., a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 9001 Quality Management System
EN ISO 13485 Quality System - Medical Devices

Safeskin Medical & Scientific (Thailand) Ltd
200 MOO 8 KANCHANAVANICH ROAD
PRIK SADA O
Amphur Sadad
SONGKHLA 90120
THAILAND

Order Information

Account Number:	101195
Synergy Health Sales Part Reference:	1134581
Customer Reference Number:	4027018042
Product Description:	HALYARD* PUREZERO* HG3 STERILE LIGHT BLUE NITRILE GLOVES - PAIRS
Validation Reference:	0.0968 Rev01
Quantity Received:	1137
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	50.0
Customer Unit Lot/Batch Number:	SEE BELOW

Processing Site: Plot 203, Kuala Ketil Industrial Park, Kuala Ketil, 09300 Phone No: +60(0)44152111

Registered Office: 170-09-01, Livingston Tower, Jalan Argyll, 10050 George Town, Pulau Pinang, MALAYSIA

VAT Number: 000859889664

Page 1 of 3


NORAZWIN BT. YUSUF
QA Executive
Synergy Sterilisation (M) Sdn. Bhd.
+60(0)44152111



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 05-Jul-2021

MY03S12583214-1-1

This is to certify that Synergy Sterilisation (M) Sdn. Bhd., a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 9001 Quality Management System
EN ISO 13485 Quality System - Medical Devices

Safeskin Medical & Scientific (Thailand) Ltd
200 MOO 8 KANCHANAVANICH ROAD
PRIK SADA O
Amphur Sadad
SONGKHLA 90120
THAILAND

Other Process Details:

HALYARD* PUREZERO* HG3 Sterile Light Blue
Nitrile Gloves - Pairs

Catalog Numbers : CLN923260, CLN923265,
CLN923270, CLN923275, CLN923280,
CLN923285, CLN923290, CLN923210.

OMH Code : 48172-00, 48173-00, 48174-00,
48175-00, 48176-00, 48177-00, 48178-00,
48179-00

Catalog Number(s)	Lot No./Batch No.	Quantity
CLN923260 48172-00	NIL/SM117026X	4
CLN923260 48172-00	NIL/SM116526X	17
CLN923265 48173-00	NIL/SM117226X	8
CLN923265 48173-00	NIL/SM116926X	15
CLN923270 48174-00	NIL/SM117326X	36
CLN923270 48174-00	NIL/SM117226X	59
CLN923270 48174-00	NIL/SM116726X	72
CLN923275 48175-00	NIL/SM117226X	69
CLN923275 48175-00	NIL/SM117026X	88
CLN923275 48175-00	NIL/SM116926X	17
CLN923275 48175-00	NIL/SM116826X	48
CLN923275 48175-00	NIL/SM116626X	72
CLN923280 48176-00	NIL/SM117226X	66
CLN923280 48176-00	NIL/SM117126X	107
CLN923280 48176-00	NIL/SM117026X	116

Processing Site: Plot 203, Kuala Ketil Industrial Park, Kuala Ketil, 09300 Phone No: +60(0)44152111

Registered Office: 170-09-01, Livingston Tower, Jalan Argyll, 10050 George Town, Pulau Pinang, MALAYSIA

VAT Number: 000859889664

Page 2 of 3

66/07/2021
NOR AZWIN BT. YUSUF
QA Executive
Synergy Sterilisation (M) Sdn. Bhd.
+60(0)44152111



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 05-Jul-2021

MY03S12583214-1-1

This is to certify that Synergy Sterilisation (M) Sdn. Bhd., a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 9001 Quality Management System
EN ISO 13485 Quality System - Medical Devices

Safeskin Medical & Scientific (Thailand) Ltd
200 MOO 8 KANCHANAVANICH ROAD
PRIK SADA O
Amphur Sadad
SONGKHLA 90120
THAILAND

CLN923280 48176-00	NIL/SM116826X	32
CLN923285 48177-00	NIL/SM117326X	20
CLN923285 48177-00	NIL/SM117226X	7
CLN923285 48177-00	NIL/SM117126X	162
CLN923290 48178-00	NIL/SM117126X	61
CLN923210 48179-00	NIL/SM117126X	32
CLN923210 48179-00	NIL/SM116726X	29

Irradiation Data

Date and Time of Irradiation:	03-Jul-2021 11:43
Reference Dose Range kGy:	32.6 - 33.5
Calculated Minimum Dose kGy:	30.2
Calculated Maximum Dose kGy:	36.5

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company

Processing Site: Plot 203, Kuala Ketil Industrial Park, Kuala Ketil, 09300 Phone No: +60(0)44152111

Registered Office: 170-09-01, Livingston Tower, Jalan Argyll, 10050 George Town, Pulau Pinang, MALAYSIA

VAT Number: 000859889664

Page 3 of 3

06/07/2021
NOR AZWIN BT. YUSUF
QA Executive
Synergy Sterilisation (M) Sdn. Bhd.
+60(0)44152111