

Certificate of Irradiation

Date Issued : **4-Mar-2022****STT20220483**

This is to certify the STERIS TOMOE (THAILAND) LTD. has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System – Medical Devices

SAFESKIN MEDICAL & SCIENTIFIC (THAILAND) LTD.

200 MOO 8, KANJANAVANICH ROAD
TAMBOL PRIK, AMPHUR SADA O
SONGKHLA
90120
THAILAND

Order Information

Account Number :	327776																																				
STERIS TOMOE Sales Part Reference :	Process Specification no. STT-SP-SPG-112																																				
Customer Reference Number :	4027019377																																				
Product Description :	HALYARD* PUREZERO* HG3 Sterile Light Blue Nitrile Gloves - Pairs																																				
Validation Reference :	PQ Report no. CHONII-PQ0017																																				
Quantity Received :	738																																				
Customer Minimum Specification kGy :	25.0																																				
Customer Maximum Specification kGy :	50.0																																				
Customer Unit Lot/Batch Number :	SEE BELOW																																				
Other Process Details :	HALYARD* PUREZERO* HG3 Sterile Light Blue Nitrile Gloves - Pairs Catalog Numbers : CLN923260, CLN923265, CLN923270, CLN923275, CLN923280, CLN923285, CLN923290,CLN923210 OMH Code : 48172-00, 48173-00, 48174-00, 48175-00, 48176-00, 48177-00, 48178-00, 48179-00																																				
	<table><tr><th>Catalog Number(s)</th><th>Lot No./Batch No.</th><th>Quantity</th></tr><tr><td>CLN923285 48177-00</td><td>NIL/SM205426X</td><td>18</td></tr><tr><td>CLN923285 48177-00</td><td>NIL/SM205326X</td><td>95</td></tr><tr><td>CLN923285 48177-00</td><td>NIL/SM205126X</td><td>131</td></tr><tr><td>CLN923285 48177-00</td><td>NIL/SM205026X</td><td>100</td></tr><tr><td>CLN923285 48177-00</td><td>NIL/SM204726V</td><td>45</td></tr><tr><td>CLN923285 48177-00</td><td>NIL/SM204526V</td><td>18</td></tr><tr><td>CLN923285 48177-00</td><td>NIL/SM204426X</td><td>42</td></tr><tr><td>CLN923290 48178-00</td><td>NIL/SM205426X</td><td>41</td></tr><tr><td>CLN923290 48178-00</td><td>NIL/SM205126X</td><td>195</td></tr><tr><td>CLN923290 48178-00</td><td>NIL/SM204526X</td><td>19</td></tr><tr><td>CLN923290 48178-00</td><td>NIL/SM204426X</td><td>34</td></tr></table>	Catalog Number(s)	Lot No./Batch No.	Quantity	CLN923285 48177-00	NIL/SM205426X	18	CLN923285 48177-00	NIL/SM205326X	95	CLN923285 48177-00	NIL/SM205126X	131	CLN923285 48177-00	NIL/SM205026X	100	CLN923285 48177-00	NIL/SM204726V	45	CLN923285 48177-00	NIL/SM204526V	18	CLN923285 48177-00	NIL/SM204426X	42	CLN923290 48178-00	NIL/SM205426X	41	CLN923290 48178-00	NIL/SM205126X	195	CLN923290 48178-00	NIL/SM204526X	19	CLN923290 48178-00	NIL/SM204426X	34
Catalog Number(s)	Lot No./Batch No.	Quantity																																			
CLN923285 48177-00	NIL/SM205426X	18																																			
CLN923285 48177-00	NIL/SM205326X	95																																			
CLN923285 48177-00	NIL/SM205126X	131																																			
CLN923285 48177-00	NIL/SM205026X	100																																			
CLN923285 48177-00	NIL/SM204726V	45																																			
CLN923285 48177-00	NIL/SM204526V	18																																			
CLN923285 48177-00	NIL/SM204426X	42																																			
CLN923290 48178-00	NIL/SM205426X	41																																			
CLN923290 48178-00	NIL/SM205126X	195																																			
CLN923290 48178-00	NIL/SM204526X	19																																			
CLN923290 48178-00	NIL/SM204426X	34																																			



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued : **4-Mar-2022**

STT20220483

This is to certify the STERIS TOMOE (THAILAND) LTD. has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System – Medical Devices

SAFESKIN MEDICAL & SCIENTIFIC (THAILAND) LTD.

200 MOO 8, KANJANAVANICH ROAD
TAMBOL PRIK, AMPHUR SADAO
SONGKHLA
90120
THAILAND

Irradiation Data

Date and Time of Irradiation :	2-March-2022
Reference Dose Range kGy :	29.4 - 33.3
Calculated Minimum Dose kGy :	25.7
Calculated Maximum Dose kGy :	42.7

Irradiation Release Authorised By STERIS TOMOE (THAILAND) LTD., a STERIS Company

Rungtina S. 04 Mar. 2022

Paradee Thammajanya
QA Manager