



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 27-May-2021

MY03S12571209-1-1

This is to certify that Synergy Sterilisation (M) Sdn. Bhd., a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 9001 Quality Management System
EN ISO 13485 Quality System - Medical Devices

Safeskin Medical & Scientific (Thailand) Ltd
200 MOO 8 KANCHANAVANICH ROAD
PRIK SADA O
Amphur Sadad
SONGKHLA 90120
THAILAND

Order Information

Account Number:	101195
Synergy Health Sales Part Reference:	1132653
Customer Reference Number:	4027017816
Product Description:	HALYARD* PUREZERO* HG3 STERILE LIGHT BLUE NITRILE GLOVES - PAIRS
Validation Reference:	0.0697 Rev02
Quantity Received:	1135
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	50.0
Customer Unit Lot/Batch Number:	SEE BELOW

Processing Site: Plot 203, Kuala Ketil Industrial Park, Kuala Ketil, 09300 Phone No: +60(0)44152111

Registered Office: 170-09-01, Livingston Tower, Jalan Argyll, 10050 George Town, Pulau Pinang, MALAYSIA

VAT Number: 000859889664

Page 1 of 3

27/05/2021

NOR AZWIN BT. YUSUF
QA Executive
Synergy Sterilisation (M) Sdn. Bhd.
+60(0)44152111



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 27-May-2021

MY03S12571209-1-1

This is to certify that Synergy Sterilisation (M) Sdn. Bhd., a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 9001 Quality Management System
EN ISO 13485 Quality System - Medical Devices

Safeskin Medical & Scientific (Thailand) Ltd
200 MOO 8 KANCHANAVANICH ROAD
PRIK SADA O
Amphur Sadad
SONGKHLA 90120
THAILAND

Other Process Details:

HALYARD* PUREZERO* HG3 Sterile Light Blue
Nitrile Gloves - Pairs
Catalog Numbers : CLN923260, CLN923265,
CLN923270, CLN923275, CLN923280,
CLN923285, CLN923290, CLN923210.

OMH Code : 48172-00, 48173-00, 48174-00,
48175-00, 48176-00, 48177-00, 48178-00,
48179-00

Catalog Number(s)	Lot No./Batch No.	Quantity
CLN923265 48173-00	NIL/SM113326X	52
CLN923265 48173-00	NIL/SM112626X	16
CLN923270 48174-00	NIL/SM113126X	97
CLN923270 48174-00	NIL/SM112926X	23
CLN923275 48175-00	NIL/SM113126X	213
CLN923275 48175-00	NIL/SM112926X	12
CLN923280 48176-00	NIL/SM113226X	180
CLN923280 48176-00	NIL/SM113126X	21

Processing Site: Plot 203, Kuala Ketil Industrial Park, Kuala Ketil, 09300 Phone No: +60(0)44152111

Registered Office: 170-09-01, Livingston Tower, Jalan Argyll, 10050 George Town, Pulau Pinang, MALAYSIA

VAT Number: 000859889664

Page 2 of 3

27/05/2021
NOR AZWIN BT. YUSUF
QA Executive
Synergy Sterilisation (M) Sdn. Bhd.
+60(0)44152111



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 27-May-2021

MY03S12571209-1-1

This is to certify that Synergy Sterilisation (M) Sdn. Bhd., a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 9001 Quality Management System
EN ISO 13485 Quality System - Medical Devices

Safeskin Medical & Scientific (Thailand) Ltd
200 MOO 8 KANCHANAVANICH ROAD
PRIK SADA O
Amphur Sadad
SONGKHLA 90120
THAILAND

	CLN923280 48176-00 NIL/SM113026X	144
	CLN923280 48176-00 NIL/SM112926X	26
	CLN923285 48177-00 NIL/SM113426X	36
	CLN923285 48177-00 NIL/SM113326X	183
	CLN923285 48177-00 NIL/SM112926X	49
	CLN923285 48177-00 NIL/SM112626X	2
	CLN923290 48178-00 NIL/SM113426X	81

Irradiation Data

Date and Time of Irradiation:	26-May-2021 05:34
Reference Dose Range kGy:	31.2 - 35.7
Calculated Minimum Dose kGy:	25.5
Calculated Maximum Dose kGy:	39.4

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company

Processing Site: Plot 203, Kuala Ketil Industrial Park, Kuala Ketil, 09300 Phone No: +60(0)44152111

Registered Office: 170-09-01, Livingston Tower, Jalan Argyll, 10050 George Town, Pulau Pinang, MALAYSIA

VAT Number: 000859889664

Page 3 of 3

27/05/2021

NOR AZWIN BT. YUSUF
QA Executive
Synergy Sterilisation (M) Sdn. Bhd.
+60(0)44152111