



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued : **27-Apr-2022**

STT20220929

This is to certify the STERIS TOMOE (THAILAND) LTD. has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1

Sterilisation of Health Care Products

EN ISO 13485

Quality System – Medical Devices

SAFESKIN MEDICAL & SCIENTIFIC (THAILAND) LTD.

200 MOO 8, KANJANAVANICH ROAD

TAMBOL PRIK, AMPHUR SADAO

SONGKHLA

90120

THAILAND

Order Information

Account Number :	327776															
STERIS TOMOE Sales Part Reference :	Process Specification no. STT-SP-SPG-112															
Customer Reference Number :	4027019687															
Product Description :	HALYARD* PUREZERO* HG3 Sterile Light Blue Nitrile Gloves - Pairs															
Validation Reference :	PQ Report no. CHONII-PQ0017															
Quantity Received :	232															
Customer Minimum Specification kGy :	25.0															
Customer Maximum Specification kGy :	50.0															
Customer Unit Lot/Batch Number :	SEE BELOW															
Other Process Details :	HALYARD* PUREZERO* HG3 Sterile Light Blue Nitrile Gloves - Pairs Catalog Numbers : CLN923260, CLN923265, CLN923270, CLN923275,CLN923280,CLN923285,CLN923290, CLN923210. OMH Code : 48172-00, 48173-00, 48174-00, 48175-00, 48176-00, 48177-00, 48178-00, 48179-00															
	<table><tr><th>Catalog Number(s)</th><th>Lot No./Batch No.</th><th>Quantity</th></tr><tr><td>CLN923260 48172-00</td><td>NIL/SM209026X</td><td>25</td></tr><tr><td>CLN923270 48174-00</td><td>NIL/SM209126X</td><td>126</td></tr><tr><td>CLN923280 48176-00</td><td>NIL/SM210726X</td><td>36</td></tr><tr><td>CLN923210 48179-00</td><td>NIL/SM209126X</td><td>45</td></tr></table>	Catalog Number(s)	Lot No./Batch No.	Quantity	CLN923260 48172-00	NIL/SM209026X	25	CLN923270 48174-00	NIL/SM209126X	126	CLN923280 48176-00	NIL/SM210726X	36	CLN923210 48179-00	NIL/SM209126X	45
Catalog Number(s)	Lot No./Batch No.	Quantity														
CLN923260 48172-00	NIL/SM209026X	25														
CLN923270 48174-00	NIL/SM209126X	126														
CLN923280 48176-00	NIL/SM210726X	36														
CLN923210 48179-00	NIL/SM209126X	45														



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued : 27-Apr-2022

STT20220929

This is to certify the STERIS TOMOE (THAILAND) LTD. has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System – Medical Devices

SAFESKIN MEDICAL & SCIENTIFIC (THAILAND) LTD.
200 MOO 8, KANJANAVANICH ROAD
TAMBOL PRIK, AMPHUR SADA O
SONGKHLA
90120
THAILAND

Irradiation Data

Date and Time of Irradiation :	26-Apr-2022
Reference Dose Range kGy :	32.1 - 32.8
Calculated Minimum Dose kGy :	28.0
Calculated Maximum Dose kGy :	42.0

Irradiation Release Authorised By STERIS TOMOE (THAILAND) LTD., a STERIS Company

Rungtina S. 28 Apr. 2022

Paradee Thammajanya
QA Manager