



DuPont Specialty Products USA, LLC
5401 Route 1
Richmond, VA 23234 USA

CERTIFICATE OF STERILITY

PRODUCT NAME: DUPONT™ TYVEK® ISOCLEAN® FROCK
PRODUCT RELEASE DATE: 4/7/2025 12:00 AM
STERILIZATION RUN NUMBER: M2500821B

EXPIRATION DATE: MAR-2030
PART NUMBER: IC270BWHMD0030WS
SIZE: MD
STERILE LOT NUMBER: AHS2513156

GAMMA IRRADIATION STERILIZATION:

VALIDATED DOSE (KGY): 30.0
MINIMUM SPECIFIED DOSE (KGY): 31.0
MAXIMUM SPECIFIED DOSE (KGY): 55.0

MINIMUM DELIVERED DOSE (KGY): 33.7
MAXIMUM DELIVERED DOSE (KGY): 50.6

DUPONT CERTIFIES THAT THE LOT NUMBER ABOVE HAS BEEN RELEASED ACCORDING TO DUPONT INTERNAL SPECIFICATIONS. THE STERILITY ASSURANCE LEVEL (SAL) 10^{-6} WAS ESTABLISHED IN ACCORDANCE TO GUIDELINES OUTLINED IN ANSI/AAMI/ISO 11137.

DO NOT USE IF PACKAGE INTEGRITY HAS BEEN COMPROMISED.

CERTIFICATES OF STERILITY CAN BE RETRIEVED AT MYCUSTOMERSERVICE.NA@DUPONT.COM OR CONTACT DUPONT CUSTOMER SERVICE AT 1-800-931-3456.

APPROVED BY:

QUALITY ASSURANCE

THIS STATEMENT IS BASED ON OUR CURRENT LEVEL OF KNOWLEDGE AND IS SUBJECT TO REVISION AS ADDITIONAL KNOWLEDGE AND EXPERIENCE ARE GAINED. IT IS THE USER'S RESPONSIBILITY TO DETERMINE THE LEVEL OF RISK AND THE PROPER PROTECTION EQUIPMENT NEEDED FOR THE USER'S PARTICULAR PURPOSE. SINCE CONDITIONS OF USE ARE OUTSIDE OF OUR CONTROL, **DUPONT MAKES NO WARRANTIES, EXPRESS OR IMPLIED, AND ASSUMES NO LIABILITY IN CONNECTION WITH ANY USE OF THIS INFORMATION.** THIS INFORMATION IS NOT INTENDED TO BE USED BY OTHERS FOR ADVERTISING, PROMOTION, OR OTHER PUBLICATION FOR COMMERCIAL PURPOSES.

©2023 DuPont. All rights reserved. DuPont™, the DuPont Oval Logo, and all trademarks and service marks denoted with ™, or ® are owned by affiliates of DuPont de Nemours, Inc. unless otherwise noted.

DPP_GL_FRM_021