

Fisher Scientific Company L.L.C. 300 Industry Drive Pittsburgh, PA 15275 Tel: 724-517-2400 Fax: 724-517-1546 fishersci.com

Catalog Number: Lot Number: Batch Number: Description: Manufacture Date: 1139415A 2406PM 2406016KPM Fisherbrand™ Powder Free Latex Gloves June 2024

Certificate of Analysis

Dimensions XL 2XL 3XL S L Size XS М 0.16 0.16 -Fingertip Thickness (mm) 0.16 --... 0.14 0.13 Palm Thickness (mm) -0.13 -... 0.11 0.11 Cuff Thickness (mm) 0.11--** -245 246 Glove Length (mm) 246 ---... 106 116 Palm Width (mm 86 ... _ --

Glove Properties

Size	XS	S	м	L	XL	2XL	3XL
Powder Content (mg/giove)	-	0.20	-	0.20	0.24	-	-
Protein Content (µg/g)	-	17.3	-	16.9	17.7	-	

State	Unaged	Aged
Tensile Strength (MPa)	24.13	20.56
Ultimate Elongation (%)	736	639

Criteria

Type of Test	AQL Level	Results
Visual Inspection – Major Defects	2,5	Pass
Visual Inspection – Minor Defects	4.0	Pass
1000mL Watertight	1.5	Pass

Test Methods

Protein Level: ASTM D5712 Powder Content: ASTM D6124 Freedom from Holes: ASTM D5151 Material Standard: ASTM D3578 Packaging: 100 gloves per dispenser

This Certificate has been compiled with data that has been provided by the manufacturer of the product. The Fisher Scientific channel has not independently verified such data. This document has been produced electronically and is valid without a signature.